

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011121\
 Data File : PL064043.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jan 2021 18:13
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 12 02:06:02 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122220.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 22 14:02:42 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.435	4.106	137.1E6	121.4E6	45.728	47.564
2)	SA Decachlor...	8.168	9.264	122.7E6	102.3E6	44.522	51.888
Target Compounds							
2)	A alpha-BHC	3.868	4.508	232.4E6	178.0E6	46.547	49.399
3)	MA gamma-BHC...	4.149	4.799	214.7E6	166.6E6	46.978	48.824
4)	MA Heptachlor	4.448	5.312	202.2E6	167.5E6	48.007	52.454
5)	MB Aldrin	4.694	5.621	185.0E6	145.9E6	46.314	50.972
6)	B beta-BHC	4.397	4.975	84067718	78460172	43.496	44.698
7)	B delta-BHC	4.600	5.196	193.3E6	141.4E6	42.653	43.015
8)	B Heptachlo...	5.135	6.015	174.0E6	135.9E6	46.325	48.987
9)	A Endosulfan I	5.471	6.378	158.5E6	124.5E6	45.767	49.501
10)	B gamma-Chl...	5.361	6.251	174.8E6	138.5E6	45.975	50.052
11)	B alpha-Chl...	5.419	6.324	172.9E6	138.9E6	46.099	50.336
12)	B 4,4'-DDE	5.586	6.480	159.0E6	121.5E6	46.250	49.211
13)	MA Dieldrin	5.713	6.638	168.4E6	127.5E6	46.463	48.930
14)	MA Endrin	5.968	6.860	153.1E6	103.5E6	52.525	53.314
15)	B Endosulfa...	6.240	7.074	148.9E6	110.4E6	45.175	46.834
16)	A 4,4'-DDD	6.099	6.978	131.8E6	97997417	45.058	46.352
17)	MA 4,4'-DDT	6.338	7.278	137.1E6	100.8E6	51.737	51.791
18)	B Endrin al...	6.408	7.199	123.6E6	94173546	44.675	46.713
19)	B Endosulfa...	6.620	7.422	141.6E6	101.1E6	44.961	44.954
20)	A Methoxychlor	6.886	7.738	71580188	57695172	54.484	54.685
21)	B Endrin ke...	7.108	7.902	160.2E6	109.2E6	46.090	44.265
22)	Mirex	7.294	8.354	119.5E6	87442922	47.600	48.476

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011121\
Data File : PL064043.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jan 2021 18:13
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 12 02:06:02 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122220.M
Quant Title : GC Extractables
QLast Update : Tue Dec 22 14:02:42 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

