

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011221\
 Data File : PL064054.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jan 2021 15:05
 Operator : AR\AJ
 Sample : PSTDICC075
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 13 03:40:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011221.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 13 03:39:52 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.435	4.106	217.1E6	188.4E6	77.902	79.207
2)	SA Decachlor...	8.169	9.266	207.3E6	162.6E6	75.683	74.051
Target Compounds							
2)	A alpha-BHC	3.869	4.508	377.0E6	284.5E6	80.403	79.412
3)	MA gamma-BHC...	4.149	4.799	348.5E6	269.8E6	79.110	79.007
4)	MA Heptachlor	4.449	5.313	323.2E6	261.8E6	78.448	77.375
5)	MB Aldrin	4.695	5.623	300.4E6	231.5E6	80.076	79.181
6)	B beta-BHC	4.398	4.975	138.0E6	119.1E6	75.873	75.299
7)	B delta-BHC	4.600	5.197	343.6E6	242.7E6	79.968	82.490
8)	B Heptachlo...	5.135	6.016	274.6E6	192.6E6	78.019	79.926
9)	A Endosulfan I	5.471	6.379	256.6E6	199.9E6	77.853	77.174
10)	B gamma-Chl...	5.362	6.252	283.0E6	221.2E6	78.772	77.874
11)	B alpha-Chl...	5.420	6.325	278.3E6	222.1E6	78.476	77.583
12)	B 4,4'-DDE	5.587	6.481	263.1E6	202.1E6	79.680	79.126
13)	MA Dieldrin	5.713	6.639	276.1E6	208.9E6	79.222	77.797
14)	MA Endrin	5.968	6.861	250.4E6	170.9E6	79.369	76.713
15)	B Endosulfa...	6.240	7.074	244.4E6	179.6E6	77.729	76.424
16)	A 4,4'-DDD	6.100	6.979	216.9E6	163.4E6	79.207	77.671
17)	MA 4,4'-DDT	6.339	7.279	231.2E6	169.8E6	86.100	77.251
18)	B Endrin al...	6.409	7.200	204.0E6	153.7E6	76.002	75.326
19)	B Endosulfa...	6.621	7.423	240.9E6	173.5E6	76.395	75.382
20)	A Methoxychlor	6.886	7.739	116.4E6	96160445	75.461	75.324
21)	B Endrin ke...	7.108	7.902	258.3E6	182.6E6	77.760	80.961
22)	Mirex	7.294	8.355	186.2E6	136.4E6	74.771	73.277

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011221\
Data File : PL064054.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jan 2021 15:05
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 13 03:40:43 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011221.M
Quant Title : GC Extractables
QLast Update : Wed Jan 13 03:39:52 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

