

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011320\
 Data File : PL055695.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jan 2020 13:26
 Operator : AJ\MA
 Sample : L1014-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 OR-03-011020

Manual Integrations
 APPROVED

mohammad
 1/14/2020 12:10:28 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 13 16:24:10 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011020.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 10 00:58:33 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.492	3.967	173.6E6	34873267	18.784	13.164 #
28)	SA Decachlor...	8.237	8.955	137.6E6	33585714	15.673	16.154
Target Compounds							
10)	B gamma-Chl...	5.421	6.040	9842848	3480323	0.735m	1.236m#
11)	B alpha-Chl...	5.486	6.114	12666258	7146318	0.965m	2.543 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011320\
Data File : PL055695.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jan 2020 13:26
Operator : AJ\MA
Sample : L1014-01
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
ECD_L
ClientSampled :
OR-03-011020


Manual Integrations
APPROVED

mohammad
1/14/2020 12:10:28 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 13 16:24:10 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011020.M
Quant Title : GC Extractables
QLast Update : Fri Jan 10 00:58:33 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.492 min
Delta R.T.: 0.000 min
Response: 173576046
Conc: 18.78 ng/ml

Instrument :
ECD_L
ClientSampleId :
OR-03-011020


Manual Integrations
APPROVED

mohammad
1/14/2020 12:10:28 PM




#1 Tetrachloro-m-xylene

R.T.: 3.967 min
Delta R.T.: -0.003 min
Response: 34873267
Conc: 13.16 ng/ml




#10 gamma-Chlordane

R.T.: 5.421 min
Delta R.T.: -0.010 min
Response: 9842848
Conc: 0.74 ng/ml m



#10 gamma-Chlordane

R.T.: 6.040 min
Delta R.T.: -0.007 min
Response: 3480323
Conc: 1.24 ng/ml m




#11 alpha-Chlordane

R.T.: 5.486 min
Delta R.T.: -0.003 min
Response: 12666258
Conc: 0.96 ng/ml m

Instrument :
ECD_L
ClientSampleId :
OR-03-011020


Manual Integrations
APPROVED

mohammad
1/14/2020 12:10:28 PM




#11 alpha-Chlordane

R.T.: 6.114 min
Delta R.T.: -0.004 min
Response: 7146318
Conc: 2.54 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.237 min
Delta R.T.: 0.000 min
Response: 137636377
Conc: 15.67 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.955 min
Delta R.T.: -0.008 min
Response: 33585714
Conc: 16.15 ng/ml