

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011321\
 Data File : PL064104.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jan 2021 19:15
 Operator : AR\AJ
 Sample : 50 PPB SPIKE TEST
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 50 PPB SPIKE TEST

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 13 20:37:23 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011221.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 13 04:58:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds						
2) A	alpha-BHC	3.869	4.509	394.8E6	290.0E6	79.369
3) MA	gamma-BHC...	4.150	4.801	346.9E6	267.1E6	74.812
4) MA	Heptachlor	4.449	5.313	308.3E6	253.5E6	70.369
5) MB	Aldrin	4.695	5.624	299.6E6	234.1E6	74.759
6) B	beta-BHC	4.398	4.977	137.0E6	129.3E6	75.132
7) B	delta-BHC	4.601	5.198	324.7E6	245.8E6	70.110
8) B	Heptachlo...	5.135	6.017	280.4E6	217.2E6	73.984
9) A	Endosulfan I	5.471	6.380	261.5E6	196.9E6	74.033
10) B	gamma-Chl...	5.361	6.253	284.7E6	217.6E6	73.486
11) B	alpha-Chl...	5.420	6.326	279.5E6	215.7E6	73.052
12) B	4,4'-DDE	5.586	6.482	264.8E6	198.2E6	75.173
13) MA	Dieldrin	5.713	6.641	276.5E6	202.8E6	74.345
14) MA	Endrin	5.968	6.862	178.0E6	121.5E6	51.247
15) B	Endosulfa...	6.241	7.075	243.7E6	176.6E6	71.894
16) A	4,4'-DDD	6.099	6.980	240.9E6	177.8E6	82.602
17) MA	4,4'-DDT	6.339	7.280	171.6E6	126.5E6	57.490
18) B	Endrin al...	6.408	7.201	227.4E6	163.8E6	79.278
19) B	Endosulfa...	6.621	7.424	237.0E6	165.0E6	69.871
20) A	Methoxychlor	6.886	7.739	89583362	70894153	53.606
21) B	Endrin ke...	7.108	7.903	284.7E6	189.3E6	79.266
22)	Mirex	7.294	8.356	191.3E6	136.7E6	70.302

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011321\
Data File : PL064104.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jan 2021 19:15
Operator : AR\AJ
Sample : 50 PPB SPIKE TEST
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
50 PPB SPIKE TEST

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 13 20:37:23 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011221.M
Quant Title : GC Extractables
QLast Update : Wed Jan 13 04:58:17 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

