

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011323\
 Data File : PL080297.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jan 2023 14:19
 Operator : AR\AJ
 Sample : 01122-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PAT-03-011123

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 12 20:36:13 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011223.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 12 01:47:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.356	2.689	18526244	22412083	9.274m	10.021
28) SA Decachlor...	8.795	7.842	19279745	25076479	13.067	11.156
Target Compounds						
12) B 4,4'-DDE	5.994	5.150	2538932	2406135	1.309m	1.025m
17) MA 4,4'-DDT	6.818	5.956	5088644	7020978	2.885m	3.332



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011223\
Data File : PL080297.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jan 2023 14:19
Operator : AR\AJ
Sample : 01122-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PAT-03-011123

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 12 20:36:13 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011223.M
Quant Title : GC Extractables
QLast Update : Thu Jan 12 01:47:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.356 min
Delta R.T.: 0.004 min
Response: 18526244
Conc: 9.27 ng/ml

Instrument :
ECD_L
ClientSampleId :
PAT-03-011123




#1 Tetrachloro-m-xylene

R.T.: 2.689 min
Delta R.T.: -0.002 min
Response: 22412083
Conc: 10.02 ng/ml




#12 4,4'-DDE

R.T.: 5.994 min
Delta R.T.: 0.008 min
Response: 2538932
Conc: 1.31 ng/ml m



#12 4,4'-DDE


R.T.: 5.150 min
Delta R.T.: -0.002 min
Response: 2406135
Conc: 1.02 ng/ml m



#17 4,4'-DDT


R.T.: 6.818 min
Delta R.T.: 0.005 min
Response: 5088644
Conc: 2.89 ng/ml

Instrument :
ECD_L
ClientSampleId :
PAT-03-011123




#17 4,4'-DDT

R.T.: 5.956 min
Delta R.T.: 0.000 min
Response: 7020978
Conc: 3.33 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.795 min
Delta R.T.: 0.007 min
Response: 19279745
Conc: 13.07 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.842 min
Delta R.T.: 0.001 min
Response: 25076479
Conc: 11.16 ng/ml