

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011419\
 Data File : PL043862.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jan 2019 12:51
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 14 21:34:45 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011119.M
 Quant Title : GC Extractables
 QLast Update : Sat Jan 12 01:24:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.525	4.007	262.7E6	99599455	53.518	51.107
28)	SA Decachlor...	8.249	9.078	295.5E6	80832127	52.287	50.884
Target Compounds							
2)	A alpha-BHC	3.964	4.398	429.6E6	138.2E6	55.879	51.238
3)	MA gamma-BHC...	4.247	4.683	388.5E6	127.4E6	55.693	50.161
4)	MA Heptachlor	4.542	5.193	384.2E6	121.1E6	55.190	50.092
5)	MB Aldrin	4.789	5.498	383.3E6	118.9E6	55.989	50.092
6)	B beta-BHC	4.499	4.854	159.1E6	56512729	53.440	50.976
7)	B delta-BHC	4.702	5.071	366.0E6	111.2E6	56.841	56.592
8)	B Heptachlor...	5.231	5.885	351.3E6	110.2E6	55.243	51.049
9)	A Endosulfan I	5.569	6.241	324.7E6	101.4E6	55.642	49.178
10)	B gamma-Chl...	5.457	6.119	356.1E6	111.3E6	55.466	49.914
11)	B alpha-Chl...	5.514	6.191	349.2E6	109.8E6	55.263	49.550
12)	B 4,4'-DDE	5.678	6.348	312.5E6	94023760	57.683	51.137
13)	MA Dieldrin	5.811	6.498	348.3E6	105.3E6	55.916	49.345
14)	MA Endrin	6.068	6.716	279.2E6	80290324	51.467	45.779
15)	B Endosulfa...	6.341	6.925	289.8E6	90532624	55.707	48.882
16)	A 4,4'-DDD	6.193	6.838	253.1E6	76043649	58.501	50.325
17)	MA 4,4'-DDT	6.430	7.136	250.5E6	75845442	54.910	45.988
18)	B Endrin al...	6.508	7.049	244.3E6	77589570	54.769	48.961
19)	B Endosulfa...	6.720	7.271	296.8E6	91690767	55.096	48.627
20)	A Methoxychlor	6.973	7.594	130.6E6	40974367	51.098	46.423
21)	B Endrin ke...	7.209	7.742	337.3E6	95916069	56.010	49.424

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011419\
Data File : PL043862.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jan 2019 12:51
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 14 21:34:45 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011119.M
Quant Title : GC Extractables
QLast Update : Sat Jan 12 01:24:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

