

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011420\
 Data File : PL055728.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jan 2020 13:02
 Operator : AJ\MA
 Sample : L1004-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 BU-RO-258

Manual Integrations
 APPROVED

mohammad
 1/15/2020 4:13:13 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 14 22:05:06 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011020.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 10 00:58:33 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.490	3.967	125.1E6	37521333	13.541	14.164
28)	SA Decachlor...	8.235	8.955	149.1E6	35567641	16.980	17.108
Target Compounds							
12)	B 4,4'-DDE	5.648	6.266	10659171	2511889	0.857m	0.945m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011420\
Data File : PL055728.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jan 2020 13:02
Operator : AJ\MA
Sample : L1004-01
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
BU-RO-258


Manual Integrations
APPROVED

mohammad
1/15/2020 4:13:13 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 14 22:05:06 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011020.M
Quant Title : GC Extractables
QLast Update : Fri Jan 10 00:58:33 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.490 min
Delta R.T.: -0.002 min
Response: 125123869
Conc: 13.54 ng/ml

Instrument :
ECD_L
ClientSampleId :
BU-RO-258


Manual Integrations
APPROVED

mohammad
1/15/2020 4:13:13 PM




#1 Tetrachloro-m-xylene

R.T.: 3.967 min
Delta R.T.: -0.003 min
Response: 37521333
Conc: 14.16 ng/ml




#12 4,4'-DDE

R.T.: 5.648 min
Delta R.T.: -0.003 min
Response: 10659171
Conc: 0.86 ng/ml m



#12 4,4'-DDE

R.T.: 6.266 min
Delta R.T.: -0.007 min
Response: 2511889
Conc: 0.95 ng/ml m




#28 Decachlorobiphenyl

R.T.: 8.235 min
Delta R.T.: -0.002 min
Response: 149115673
Conc: 16.98 ng/ml

Instrument :
ECD_L
ClientSampleId :
BU-RO-258

Manual Integrations
APPROVED

mohammad
1/15/2020 4:13:13 PM



#28 Decachlorobiphenyl

R.T.: 8.955 min
Delta R.T.: -0.008 min
Response: 35567641
Conc: 17.11 ng/ml