

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011423\
 Data File : PL080307.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jan 2023 10:30
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 15 20:21:06 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011223.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 12 01:47:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.359	2.690	109.0E6	108.7E6	54.557	48.590
28)	SA Decachlor...	8.800	7.844	74619424	97172940	50.576	43.229
Target Compounds							
2)	A alpha-BHC	3.807	3.192	164.3E6	157.4E6	56.902	49.139
3)	MA gamma-BHC...	4.136	3.521	151.3E6	146.1E6	56.617	48.391
4)	MA Heptachlor	4.720	3.862	145.3E6	151.5E6	55.830	47.042
5)	MB Aldrin	5.058	4.143	134.1E6	142.5E6	56.412	47.885
6)	B beta-BHC	4.332	3.817	65564756	64112686	55.470	47.635
7)	B delta-BHC	4.575	4.047	146.0E6	141.4E6	55.866	48.926
8)	B Heptachlo...	5.483	4.645	121.9E6	131.3E6	54.898	46.822
9)	A Endosulfan I	5.865	5.015	111.3E6	122.3E6	54.934	46.769
10)	B gamma-Chl...	5.738	4.897	121.5E6	134.2E6	55.488	46.955
11)	B alpha-Chl...	5.817	4.961	119.8E6	133.7E6	55.221	46.738
12)	B 4,4'-DDE	5.995	5.154	107.3E6	114.9E6	55.297	48.942
13)	MA Dieldrin	6.139	5.281	114.0E6	126.9E6	55.504	46.578
14)	MA Endrin	6.367	5.555	95486149	107.2E6	52.058	45.068
15)	B Endosulfa...	6.585	5.849	95138725	103.7E6	49.324	42.771
16)	A 4,4'-DDD	6.509	5.708	90462598	95498525	57.158	51.526
17)	MA 4,4'-DDT	6.822	5.960	88499354	95490977	50.182	45.315
18)	B Endrin al...	6.714	6.030	75731156	81142073	52.495	44.568
19)	B Endosulfa...	6.948	6.253	94922389	105.0E6	54.460	45.221
20)	A Methoxychlor	7.300	6.538	45513746	52051953	51.084	43.304
21)	B Endrin ke...	7.429	6.758	102.5E6	115.3E6	55.039	45.799
22)	Mirex	7.900	6.943	75080769	96773966	53.078	42.994

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011423\
Data File : PL080307.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jan 2023 10:30
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 15 20:21:06 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011223.M
Quant Title : GC Extractables
QLast Update : Thu Jan 12 01:47:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

