

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011520\
 Data File : PL055769.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jan 2020 16:55
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

mohammad
 1/16/2020 11:30:33 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 16 00:37:55 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011020.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 10 00:58:33 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.491	3.969	508.5E6	148.3E6	55.025	55.978
28)	SA Decachlor...	8.237	8.958	460.4E6	107.5E6	52.428	51.706
Target Compounds							
2)	A alpha-BHC	3.929	4.352	871.9E6	227.5E6	55.160	57.383
3)	MA gamma-BHC...	4.212	4.632	815.9E6	211.6E6	55.235	56.928
4)	MA Heptachlor	4.513	5.132	786.0E6	180.3E6	54.946	56.816
5)	MB Aldrin	4.762	5.432	822.6E6	177.4E6	58.638	56.887
6)	B beta-BHC	4.461	4.800	332.7E6	83841759	54.432	54.916
7)	B delta-BHC	4.666	5.014	790.9E6	188.7E6	55.329	61.492
8)	B Heptachlo...	5.203	5.813	729.0E6	159.9E6	54.920	54.409
9)	A Endosulfan I	5.543	6.165	666.2E6	147.9E6	53.922	55.876
10)	B gamma-Chl...	5.430	6.044	772.4E6	158.3E6	57.691	56.239
11)	B alpha-Chl...	5.488	6.116	733.8E6	156.9E6	55.890	55.839
12)	B 4,4'-DDE	5.650	6.270	519.5E6	148.8E6	41.776m	55.996 #
13)	MA Dieldrin	5.785	6.420	714.3E6	155.9E6	54.495	57.003
14)	MA Endrin	6.042	6.635	625.6E6	127.5E6	54.170	56.554
15)	B Endosulfa...	6.314	6.842	624.5E6	130.5E6	57.892	54.305
16)	A 4,4'-DDD	6.164	6.755	568.5E6	123.0E6	59.990m	59.372
17)	MA 4,4'-DDT	6.406	7.051	505.2E6	106.3E6	58.831	53.931
18)	B Endrin al...	6.481	6.966	504.6E6	109.6E6	53.459	53.961
19)	B Endosulfa...	6.695	7.186	550.6E6	118.8E6	56.284	53.382
20)	A Methoxychlor	6.950	7.504	249.3E6	55617182	58.076	53.128
21)	B Endrin ke...	7.186	7.653	611.9E6	131.8E6	53.980	55.374
22)	Mirex	7.375	8.103	391.6E6	86917170	57.192	53.013

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011520\
Data File : PL055769.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jan 2020 16:55
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

mohammad
1/16/2020 11:30:33 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 16 00:37:55 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011020.M
Quant Title : GC Extractables
QLast Update : Fri Jan 10 00:58:33 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

