

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011521\
 Data File : PL064152.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jan 2021 10:07
 Operator : AR\AJ
 Sample : PB134132BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB134132BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 16 00:03:42 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011221.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 13 04:58:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.443	4.108	62680990	51410314	20.904	20.110
28)	SA Decachlor...	8.179	9.271	116.0E6	87938052	38.010	37.051
Target Compounds							
14)	MA Endrin	0.000	6.868	0	1963192	N.D.	0.860 #
15)	B Endosulfa...	0.000	7.058f	0	3318181	N.D.	1.314 #
21)	B Endrin ke...	0.000	7.910	0	140100	N.D.	0.056 #
23)	Chlordane-1	0.000	5.099f	0	2008304	N.D.	19.058 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011521\
Data File : PL064152.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jan 2021 10:07
Operator : AR\AJ
Sample : PB134132BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB134132BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 16 00:03:42 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011221.M
Quant Title : GC Extractables
QLast Update : Wed Jan 13 04:58:17 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.443 min
Delta R.T.: 0.008 min
Response: 62680990
Conc: 20.90 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB134132BL




#1 Tetrachloro-m-xylene

R.T.: 4.108 min
Delta R.T.: 0.000 min
Response: 51410314
Conc: 20.11 ng/ml




#14 Endrin

R.T.: 0.000 min
Exp R.T. : 5.969 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 6.868 min
Delta R.T.: 0.007 min
Response: 1963192
Conc: 0.86 ng/ml



#15 Endosulfan II


R.T.: 0.000 min
Exp R.T. : 6.240 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
PB134132BL




#15 Endosulfan II

R.T.: 7.058 min
Delta R.T.: -0.016 min
Response: 3318181
Conc: 1.31 ng/ml




#21 Endrin ketone

R.T.: 0.000 min
Exp R.T. : 7.108 min
Response: 0
Conc: N.D.



#21 Endrin ketone


R.T.: 7.910 min
Delta R.T.: 0.007 min
Response: 140100
Conc: 0.06 ng/ml



#23 Chlordane-1


R.T.: 0.000 min
Exp R.T.: 4.297 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
PB134132BL




#23 Chlordane-1

R.T.: 5.099 min
Delta R.T.: -0.022 min
Response: 2008304
Conc: 19.06 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.179 min
Delta R.T.: 0.011 min
Response: 116009754
Conc: 38.01 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.271 min
Delta R.T.: 0.005 min
Response: 87938052
Conc: 37.05 ng/ml