

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011620\
 Data File : PL055799.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jan 2020 16:36
 Operator : AJ\MA
 Sample : L1006-01
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 HD-01-011520

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 17 00:42:31 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011020.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 10 00:58:33 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.491	3.968	114.1E6	35244292	12.346	13.304
28)	SA Decachlor...	8.238	8.958	118.0E6	29243612	13.442	14.066
Target Compounds							
18)	B Endrin al...	6.482	0.000	23046529	0	2.442	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011620\
Data File : PL055799.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jan 2020 16:36
Operator : AJ\MA
Sample : L1006-01
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
HD-01-011520

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 17 00:42:31 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011020.M
Quant Title : GC Extractables
QLast Update : Fri Jan 10 00:58:33 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.491 min
Delta R.T.: -0.001 min
Response: 114087171
Conc: 12.35 ng/ml

Instrument :
ECD_L
ClientSampleId :
HD-01-011520




#1 Tetrachloro-m-xylene

R.T.: 3.968 min
Delta R.T.: -0.002 min
Response: 35244292
Conc: 13.30 ng/ml




#18 Endrin aldehyde

R.T.: 6.482 min
Delta R.T.: 0.000 min
Response: 23046529
Conc: 2.44 ng/ml



#18 Endrin aldehyde


R.T.: 0.000 min
Exp R.T. : 6.968 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.238 min
Delta R.T.: 0.000 min
Response: 118046614
Conc: 13.44 ng/ml

Instrument :
ECD_L
ClientSampleId :
HD-01-011520



#28 Decachlorobiphenyl

R.T.: 8.958 min
Delta R.T.: -0.005 min
Response: 29243612
Conc: 14.07 ng/ml