

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011719\
 Data File : PL044012.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jan 2019 14:09
 Operator : AJ\SJ
 Sample : K1147-04
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 06:39:00 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011719.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 18 06:10:39 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.567	4.008	52557634	21287665	9.122	9.622
28)	SA Decachlor...	8.293	9.078	72233035	26422009	10.692	12.372
Target Compounds							
8)	B Heptachlo...	0.000	5.879	0	13363599	N.D.	5.664 #
15)	B Endosulfa...	6.320f	0.000	21152961	0	3.126	N.D. #
16)	A 4,4'-DDD	0.000	6.798f	0	10845836	N.D.	5.926 #
20)	A Methoxychlor	7.002	0.000	10023565	0	3.002	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011719\
Data File : PL044012.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jan 2019 14:09
Operator : AJ\SJ
Sample : K1147-04
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TP-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 18 06:39:00 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011719.M
Quant Title : GC Extractables
QLast Update : Fri Jan 18 06:10:39 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.567 min
Delta R.T.: 0.008 min
Response: 52557634
Conc: 9.12 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-2




#1 Tetrachloro-m-xylene

R.T.: 4.008 min
Delta R.T.: 0.000 min
Response: 21287665
Conc: 9.62 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 5.270 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 5.879 min
Delta R.T.: -0.005 min
Response: 13363599
Conc: 5.66 ng/ml



#15 Endosulfan II


R.T.: 6.320 min
Delta R.T.: -0.060 min
Response: 21152961
Conc: 3.13 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-2




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 6.924 min
Response: 0
Conc: N.D.




#16 4,4'-DDD

R.T.: 0.000 min
Exp R.T. : 6.231 min
Response: 0
Conc: N.D.



#16 4,4'-DDD


R.T.: 6.798 min
Delta R.T.: -0.040 min
Response: 10845836
Conc: 5.93 ng/ml



#20 Methoxychlor


R.T.: 7.002 min
Delta R.T.: -0.007 min
Response: 10023565
Conc: 3.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP-2




#20 Methoxychlor

R.T.: 0.000 min
Exp R.T.: 7.593 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.293 min
Delta R.T.: 0.009 min
Response: 72233035
Conc: 10.69 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.078 min
Delta R.T.: 0.002 min
Response: 26422009
Conc: 12.37 ng/ml