

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011719\
 Data File : PL044030.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jan 2019 18:27
 Operator : AJ\SJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 07:07:53 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011719.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 18 06:10:39 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.570	4.007	118.2E6	47303511	20.512	21.381
28)	SA Decachlor...	8.297	9.078	132.4E6	44926649	19.595	21.037
Target Compounds							
6)	B beta-BHC	4.500f	0.000	5282745	0	1.509	N.D. #
15)	B Endosulfa...	6.324f	0.000	31402937	0	4.641	N.D. #
16)	A 4,4'-DDD	0.000	6.799f	0	9308129	N.D.	5.086 #
23)	Chlordane-1	4.500	0.000	5282745	0	20.045	N.D. #
27)	Chlordane-5	6.324	0.000	31402937	0	119.098	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011719\
Data File : PL044030.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jan 2019 18:27
Operator : AJ\SJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 18 07:07:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011719.M
Quant Title : GC Extractables
QLast Update : Fri Jan 18 06:10:39 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.570 min
Delta R.T.: 0.011 min
Response: 118186262
Conc: 20.51 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 4.007 min
Delta R.T.: 0.000 min
Response: 47303511
Conc: 21.38 ng/ml




#6 beta-BHC

R.T.: 4.500 min
Delta R.T.: -0.035 min
Response: 5282745
Conc: 1.51 ng/ml



#6 beta-BHC


R.T.: 0.000 min
Exp R.T. : 4.854 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 6.324 min
Delta R.T.: -0.057 min
Response: 31402937
Conc: 4.64 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 6.924 min
Response: 0
Conc: N.D.




#16 4,4'-DDD

R.T.: 0.000 min
Exp R.T. : 6.231 min
Response: 0
Conc: N.D.



#16 4,4'-DDD


R.T.: 6.799 min
Delta R.T.: -0.038 min
Response: 9308129
Conc: 5.09 ng/ml



#23 Chlordane-1


R.T.: 4.500 min
Delta R.T.: 0.068 min
Response: 5282745
Conc: 20.05 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#23 Chlordane-1

R.T.: 0.000 min
Exp R.T. : 5.003 min
Response: 0
Conc: N.D.




#27 Chlordane-5

R.T.: 6.324 min
Delta R.T.: -0.068 min
Response: 31402937
Conc: 119.10 ng/ml



#27 Chlordane-5


R.T.: 0.000 min
Exp R.T. : 6.995 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.297 min
Delta R.T.: 0.013 min
Response: 132377724
Conc: 19.60 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK



#28 Decachlorobiphenyl

R.T.: 9.078 min
Delta R.T.: 0.002 min
Response: 44926649
Conc: 21.04 ng/ml