

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011720\
 Data File : PL055813.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jan 2020 13:06
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

mohammad
 1/20/2020 10:38:00 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 01:14:05 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011020.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 10 00:58:33 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.492	3.965	486.8E6	139.8E6	52.683	52.754
28)	SA Decachlor...	8.237	8.956	442.3E6	99012888	50.371	47.624
Target Compounds							
2)	A alpha-BHC	3.930	4.348	837.5E6	213.5E6	52.984	53.866
3)	MA gamma-BHC...	4.214	4.628	790.1E6	197.2E6	53.491	53.038
4)	MA Heptachlor	4.515	5.128	760.4E6	172.7E6	53.160	54.417
5)	MB Aldrin	4.763	5.428	735.5E6	166.3E6	52.429	53.319
6)	B beta-BHC	4.462	4.795	313.6E6	91168013	51.315	59.715m
7)	B delta-BHC	4.667	5.011	754.2E6	162.3E6	52.762	52.864
8)	B Heptachlo...	5.205	5.809	670.4E6	138.8E6	50.505	47.247
9)	A Endosulfan I	5.543	6.162	594.8E6	138.1E6	48.140	52.154
10)	B gamma-Chl...	5.431	6.041	676.1E6	147.2E6	50.500	52.277
11)	B alpha-Chl...	5.489	6.112	645.0E6	146.2E6	49.120	51.999
12)	B 4,4'-DDE	5.652	6.267	577.6E6	140.2E6	46.454	52.769
13)	MA Dieldrin	5.786	6.416	622.3E6	143.0E6	47.473	52.308
14)	MA Endrin	6.043	6.632	532.4E6	116.8E6	46.105	51.782
15)	B Endosulfa...	6.315	6.839	497.9E6	122.9E6	46.153	51.158
16)	A 4,4'-DDD	6.167	6.753	461.5E6	110.7E6	48.700	53.433
17)	MA 4,4'-DDT	6.407	7.048	482.8E6	106.1E6	56.218	53.835
18)	B Endrin al...	6.483	6.962	470.6E6	103.0E6	49.861	50.716
19)	B Endosulfa...	6.696	7.184	467.5E6	113.0E6	47.789	50.791
20)	A Methoxychlor	6.950	7.502	254.3E6	52909170	59.248	50.541
21)	B Endrin ke...	7.186	7.650	578.4E6	121.0E6	51.023	50.806
22)	Mirex	7.375	8.100	321.0E6	79689479	46.883	48.605

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011720\
Data File : PL055813.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jan 2020 13:06
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

mohammad
1/20/2020 10:38:00 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 18 01:14:05 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011020.M
Quant Title : GC Extractables
QLast Update : Fri Jan 10 00:58:33 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

