

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011820\
 Data File : PL055843.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jan 2020 07:09
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 20 00:31:28 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011020.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 10 00:58:33 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.490	3.967	458.7E6	135.5E6	49.638	51.153
2)	SA Decachlor...	8.234	8.954	459.0E6	100.9E6	52.270	48.542
Target Compounds							
2)	A alpha-BHC	3.927	4.351	784.2E6	209.7E6	49.617	52.890
3)	MA gamma-BHC...	4.210	4.630	732.8E6	193.6E6	49.608	52.070
4)	MA Heptachlor	4.511	5.130	703.3E6	163.4E6	49.163	51.511
5)	MB Aldrin	4.760	5.430	724.1E6	164.5E6	51.613	52.727
6)	B beta-BHC	4.459	4.799	300.9E6	77363151	49.234	50.673
7)	B delta-BHC	4.664	5.012	728.0E6	165.7E6	50.934	53.973
8)	B Heptachlo...	5.201	5.811	665.4E6	147.1E6	50.126	50.050
9)	A Endosulfan I	5.540	6.163	606.3E6	137.8E6	49.072	52.050
10)	B gamma-Chl...	5.427	6.042	673.6E6	146.8E6	50.318	52.143
11)	B alpha-Chl...	5.485	6.114	653.4E6	145.6E6	49.766	51.799
12)	B 4,4'-DDE	5.649	6.269	568.0E6	139.3E6	45.678	52.443
13)	MA Dieldrin	5.782	6.417	659.1E6	143.3E6	50.279	52.408
14)	MA Endrin	6.040	6.633	522.2E6	104.0E6	45.218	46.113
15)	B Endosulfa...	6.311	6.840	546.4E6	123.6E6	50.648	51.454
16)	A 4,4'-DDD	6.163	6.754	540.6E6	118.1E6	57.045	56.974
17)	MA 4,4'-DDT	6.403	7.048	426.2E6	90862874	49.626	46.083
18)	B Endrin al...	6.479	6.963	474.1E6	104.3E6	50.222	51.368
19)	B Endosulfa...	6.691	7.183	518.1E6	113.3E6	52.968	50.943
20)	A Methoxychlor	6.946	7.502	219.6E6	47779869	51.149	45.642
21)	B Endrin ke...	7.182	7.650	605.7E6	128.5E6	53.437	53.988
22)	Mirex	7.371	8.100	367.5E6	81652641	53.670	49.802

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL011820\
Data File : PL055843.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jan 2020 07:09
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 100 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 20 00:31:28 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011020.M
Quant Title : GC Extractables
QLast Update : Fri Jan 10 00:58:33 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

