

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX012020.M
 Title : GC Extractables
 Last Update : Tue Jan 21 03:01:10 2020
 Response Via : Initial Calibration

Calibration Files

500 =PL055874.D 1000 =PL055872.D 250 =PL055875.D
 750 =PL055873.D 100 =PL055876.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	9.879	9.912	9.783	9.834	9.942	9.870	E6	0.64
2)	Toxaphene-1	6.954	8.256	7.765	7.819	7.883	7.736	E4	6.17
3)	Toxaphene-2	6.153	7.352	5.836	8.165	6.039	6.709	E4	15.01
4)	Toxaphene-3	7.298	8.511	8.003	8.323	6.850	7.797	E4	9.01
5)	Toxaphene-4	1.842	2.082	2.023	2.205	2.014	2.033	E5	6.45
6)	Toxaphene-5	1.919	2.069	2.099	2.223	2.239	2.110	E5	6.16
7) SA	Decachlorobiphenyl	0.985	1.041	1.084	1.035	1.091	1.047	E7	4.09

Signal #2 Calibration Files

500 =PL055874.D 1000 =PL055872.D 250 =PL055875.D
 750 =PL055873.D 100 =PL055876.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.846	2.846	2.864	2.870	3.018	2.889	E6	2.52
2)	Toxaphene-1	1.479	1.343	1.425	1.413	1.427	1.417	E4	3.46
3)	Toxaphene-2	2.444	2.302	2.462	2.377	2.645	2.446	E4	5.22
4)	Toxaphene-3	6.280	6.141	6.479	6.283	6.333	6.303	E4	1.93
5)	Toxaphene-4	3.999	3.899	4.066	3.877	3.806	3.929	E4	2.62
6)	Toxaphene-5	3.725	3.663	3.907	3.737	3.866	3.780	E4	2.71
7) SA	Decachlorobiphenyl	2.256	2.186	2.407	2.220	2.550	2.324	E6	6.54

(#) = Out of Range