

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012020\
 Data File : PL055883.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jan 2020 23:44
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 03:06:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012020.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 02:21:03 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.490	3.966	492.3E6	144.1E6	51.362	50.978
28)	SA Decachlor...	8.233	8.952	533.7E6	114.4E6	51.047	49.919
Target Compounds							
2)	A alpha-BHC	3.927	4.350	845.8E6	221.2E6	50.971	52.150
3)	MA gamma-BHC...	4.210	4.629	797.4E6	206.0E6	52.559	51.923
4)	MA Heptachlor	4.511	5.129	772.7E6	170.9E6	51.748	50.540
5)	MB Aldrin	4.759	5.429	773.0E6	175.2E6	52.816	51.104
6)	B beta-BHC	4.459	4.798	322.1E6	87650362	50.189	55.232
7)	B delta-BHC	4.664	5.011	773.9E6	166.6E6	52.798	55.756
8)	B Heptachlo...	5.201	5.810	718.0E6	157.1E6	52.050	52.584
9)	A Endosulfan I	5.540	6.162	667.2E6	147.9E6	51.424	50.785
10)	B gamma-Chl...	5.428	6.041	739.3E6	157.0E6	52.694	50.915
11)	B alpha-Chl...	5.486	6.112	719.0E6	156.8E6	52.011	50.788
12)	B 4,4'-DDE	5.649	6.267	628.6E6	149.6E6	48.126	51.083
13)	MA Dieldrin	5.782	6.417	725.2E6	154.8E6	52.431	51.238
14)	MA Endrin	6.039	6.632	645.2E6	129.4E6	51.074	51.054
15)	B Endosulfa...	6.311	6.839	541.5E6	133.6E6	48.491	49.808
16)	A 4,4'-DDD	6.163	6.753	522.2E6	118.9E6	52.438	51.593
17)	MA 4,4'-DDT	6.403	7.047	529.1E6	114.4E6	52.280	50.653
18)	B Endrin al...	6.479	6.962	528.8E6	114.0E6	50.882	50.482
19)	B Endosulfa...	6.691	7.183	515.7E6	124.6E6	47.682	50.342
20)	A Methoxychlor	6.946	7.501	269.6E6	60287601	51.669	50.526
21)	B Endrin ke...	7.182	7.649	656.1E6	135.1E6	52.114	51.180
22)	Mirex	7.372	8.098	366.4E6	92354578	52.165	50.614

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012020\
 Data File : PL055883.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jan 2020 23:44
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampled :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 03:06:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012020.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 02:21:03 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

