

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012021\
 Data File : PL064254.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jan 2021 14:32
 Operator : AR\AJ
 Sample : M1169-01DL 5X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-DDL

Manual Integrations
 APPROVED

Ankita
 1/21/2021 10:39:36 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 20 15:51:19 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011221.M
 Quant Title : GC Extractables
 QLast Update : Mon Jan 18 13:25:12 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.434	4.107	13648943	10946185	4.552	4.282
28) SA Decachlor...	8.166	9.268	10631599	7099065	3.483	2.991
Target Compounds						
11) B alpha-Chl...	5.417	6.325	1477264	1652706	0.386m	0.543m#
12) B 4,4'-DDE	5.584	6.483	19490312	14251736	5.533	5.351
16) A 4,4'-DDD	6.097	6.981	21971461	16675956	7.533	7.523
17) MA 4,4'-DDT	6.337	7.281	179.9E6	121.4E6	60.250	53.579

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012021\
Data File : PL064254.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jan 2021 14:32
Operator : AR\AJ
Sample : M1169-01DL 5X
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TP-DDL

Manual Integrations
APPROVED

Ankita
1/21/2021 10:39:36 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 20 15:51:19 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011221.M
Quant Title : GC Extractables
QLast Update : Mon Jan 18 13:25:12 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

