

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012120\
 Data File : PL055949.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jan 2020 17:05
 Operator : AJ\MA
 Sample : PB126245BS
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB126245BS

Manual Integrations
 APPROVED

Ankita
 1/22/2020 11:00:05 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 22 01:17:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012020.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 02:21:03 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.489	3.966	198.6E6	64119475	20.721	22.687
28)	SA Decachlor...	8.233	8.951	231.7E6	51391674	22.158	22.432
Target Compounds							
2)	A alpha-BHC	3.926	4.349	787.4E6	218.0E6	47.454	51.409
3)	MA gamma-BHC...	4.209	4.628	751.2E6	200.9E6	49.516	50.637
4)	MA Heptachlor	4.511	5.128	738.2E6	174.8E6	49.438	51.697
5)	MB Aldrin	4.759	5.428	760.1E6	180.8E6	51.933	52.743
6)	B beta-BHC	4.458	4.797	322.6E6	88842197	50.262	55.983
7)	B delta-BHC	4.664	5.009	678.3E6	150.2E6	46.279	50.253m
8)	B Heptachlo...	5.201	5.809	705.0E6	149.2E6	51.107	49.944
9)	A Endosulfan I	5.539	6.162	660.3E6	148.9E6	50.887	51.148
10)	B gamma-Chl...	5.427	6.040	737.1E6	160.5E6	52.539	52.055
11)	B alpha-Chl...	5.486	6.111	697.9E6	157.2E6	50.483	50.915
12)	B 4,4'-DDE	5.648	6.266	679.3E6	153.0E6	52.010	52.249
13)	MA Dieldrin	5.782	6.415	716.2E6	158.0E6	51.776	52.274
14)	MA Endrin	6.039	6.631	624.7E6	126.2E6	49.444	49.825
15)	B Endosulfa...	6.311	6.838	591.3E6	132.7E6	52.947	49.470
16)	A 4,4'-DDD	6.163	6.751	523.2E6	122.1E6	52.540	52.966
17)	MA 4,4'-DDT	6.403	7.047	550.8E6	119.1E6	54.425	52.748
18)	B Endrin al...	6.479	6.961	504.5E6	112.8E6	48.542	49.947
19)	B Endosulfa...	6.691	7.182	569.9E6	124.8E6	52.696	50.411
20)	A Methoxychlor	6.945	7.500	261.4E6	60038610	50.092	50.317
21)	B Endrin ke...	7.179	7.648	656.9E6	139.9E6	52.182m	52.997
22)	Mirex	7.370	8.097	422.7E6	91215728	60.186	49.990

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012120\
Data File : PL055949.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2020 17:05
Operator : AJ\MA
Sample : PB126245BS
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB126245BS

Manual Integrations
APPROVED

Ankita
1/22/2020 11:00:05 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 22 01:17:42 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012020.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 02:21:03 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

