

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012121\
 Data File : PL064267.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jan 2021 19:01
 Operator : AR\AJ
 Sample : INDA101
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 INDA101

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 03:11:03 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012121CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 21 03:09:57 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.434	4.108	14339066	11760695	4.870	4.843
27)	SA Decachlor...	8.166	9.270	33108613	25022270	10.917	10.633
Target Compounds							
2)	A alpha-BHC	3.868	4.509	21151337	13749238	4.726	4.334
3)	MA gamma-BHC...	4.148	4.801	21917411	13751096	5.126	4.505
4)	MA Heptachlor	4.448	5.314	20395189	15455826	4.767	4.733
9)	A Endosulfan I	5.470	6.380	16933851	12313297	4.865	4.715
13)	MA Dieldrin	5.712	6.641	36594306	25656781	9.435	9.155
14)	MA Endrin	5.966	6.864	33648027	20630363	9.668	8.872
16)	A 4,4'-DDD	6.097	6.981	27212246	21266718	9.137	9.565
17)	MA 4,4'-DDT	6.337	7.282	27688220	18674745	9.084	8.591
20)	A Methoxychlor	6.884	7.742	82932381	62929718	50.781	48.853


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012121\
Data File : PL064267.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jan 2021 19:01
Operator : AR\AJ
Sample : INDA101
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 03:11:03 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012121CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jan 21 03:09:57 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.434 min
Delta R.T.: 0.000 min
Response: 14339066
Conc: 4.87 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDA101




#1 Tetrachloro-m-xylene

R.T.: 4.108 min
Delta R.T.: 0.000 min
Response: 11760695
Conc: 4.84 ng/ml




#2 alpha-BHC

R.T.: 3.868 min
Delta R.T.: 0.000 min
Response: 21151337
Conc: 4.73 ng/ml



#2 alpha-BHC


R.T.: 4.509 min
Delta R.T.: 0.000 min
Response: 13749238
Conc: 4.33 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.148 min
Delta R.T.: 0.000 min
Response: 21917411
Conc: 5.13 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDA101




#3 gamma-BHC (Lindane)

R.T.: 4.801 min
Delta R.T.: 0.000 min
Response: 13751096
Conc: 4.50 ng/ml




#4 Heptachlor

R.T.: 4.448 min
Delta R.T.: 0.000 min
Response: 20395189
Conc: 4.77 ng/ml



#4 Heptachlor


R.T.: 5.314 min
Delta R.T.: 0.000 min
Response: 15455826
Conc: 4.73 ng/ml



#9 Endosulfan I


R.T.: 5.470 min
Delta R.T.: 0.000 min
Response: 16933851
Conc: 4.87 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDA101




#9 Endosulfan I

R.T.: 6.380 min
Delta R.T.: 0.000 min
Response: 12313297
Conc: 4.72 ng/ml




#13 Dieldrin

R.T.: 5.712 min
Delta R.T.: 0.000 min
Response: 36594306
Conc: 9.43 ng/ml



#13 Dieldrin


R.T.: 6.641 min
Delta R.T.: 0.000 min
Response: 25656781
Conc: 9.16 ng/ml



#14 Endrin


R.T.: 5.966 min
Delta R.T.: 0.000 min
Response: 33648027
Conc: 9.67 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDA101




#14 Endrin

R.T.: 6.864 min
Delta R.T.: 0.000 min
Response: 20630363
Conc: 8.87 ng/ml




#16 4,4'-DDD

R.T.: 6.097 min
Delta R.T.: 0.000 min
Response: 27212246
Conc: 9.14 ng/ml



#16 4,4'-DDD


R.T.: 6.981 min
Delta R.T.: 0.000 min
Response: 21266718
Conc: 9.56 ng/ml



#17 4,4'-DDT


R.T.: 6.337 min
Delta R.T.: 0.000 min
Response: 27688220
Conc: 9.08 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDA101




#17 4,4'-DDT

R.T.: 7.282 min
Delta R.T.: 0.000 min
Response: 18674745
Conc: 8.59 ng/ml




#20 Methoxychlor

R.T.: 6.884 min
Delta R.T.: 0.000 min
Response: 82932381
Conc: 50.78 ng/ml



#20 Methoxychlor


R.T.: 7.742 min
Delta R.T.: 0.000 min
Response: 62929718
Conc: 48.85 ng/ml



#27 Decachlorobiphenyl

R.T.: 8.166 min
Delta R.T.: -0.001 min
Response: 33108613
Conc: 10.92 ng/ml

Instrument :
ECD_L
ClientSampleId :
INDA101



#27 Decachlorobiphenyl

R.T.: 9.270 min
Delta R.T.: 0.000 min
Response: 25022270
Conc: 10.63 ng/ml