

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012122\
 Data File : PL072152.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jan 2022 09:55
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 01/24/2022
 Supervised By : Ankita Jodhani 01/25/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 22 01:55:40 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011722.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 18 10:53:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.562	5.492	26260859	79067747	47.301	48.684
28)	SA Decachlor...	10.321	11.384	38477953	76042345	43.310	47.129
Target Compounds							
2)	A alpha-BHC	5.268	6.074	36845055	119.0E6	47.651	49.320
3)	MA gamma-BHC...	5.690	6.470	35438167	108.7E6	46.999	48.307
4)	MA Heptachlor	6.096	7.116	39622863	108.3E6	45.929	46.724
5)	MB Aldrin	6.420	7.486	36626552	104.9E6	46.384	47.451
6)	B beta-BHC	6.075	6.729	15330290	49397985	46.376	49.891
7)	B delta-BHC	6.333	6.995	37330272	109.3E6	47.126	44.415
8)	B Heptachlo...	6.997	7.953	35329374	95625846	46.411	47.758
9)	A Endosulfan I	7.399	8.358	32979225	86687074	46.403	47.891
10)	B gamma-Chl...	7.274	8.223	35981715	97058138	45.795	48.029
11)	B alpha-Chl...	7.343	8.303	35966925	94641722	45.847	47.731
12)	B 4,4'-DDE	7.558	8.494	33646790	91733555	46.603	48.504
13)	MA Dieldrin	7.685	8.647	35285072	93056806	46.574	47.367
14)	MA Endrin	7.976	8.887	31016642	75127921	44.267	45.312
15)	B Endosulfa...	8.288	9.123	33100508	79367103	44.518	46.281
16)	A 4,4'-DDD	8.144	9.034	28392814	70743730	48.434	49.479m
17)	MA 4,4'-DDT	8.404	9.350	28707738	70490763	43.230	44.599
18)	B Endrin al...	8.477	9.258	25576056	59113446	44.890	47.619
19)	B Endosulfa...	8.708	9.495	33106418	73699939	45.154	47.209
20)	A Methoxychlor	9.000	9.839	18119453	36741402	42.948	44.355
21)	B Endrin ke...	9.225	9.992	39132309	79595692	46.995	48.094
22)	Mirex	9.407	10.454	32824977	56323872	44.449	45.139

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012122\
 Data File : PL072152.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jan 2022 09:55
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 01/24/2022
 Supervised By :Ankita Jodhani 01/25/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 22 01:55:40 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011722.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 18 10:53:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

