

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012122\
 Data File : PL072158.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jan 2022 12:09
 Operator : AR\AJ
 Sample : PB142173BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB142173BL

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 01/24/2022
 Supervised By : Ankita Jodhani 01/25/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 22 01:57:39 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011722.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 18 10:53:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.555	5.491	9847927	28882340	17.738	17.783
2) SA Decachlor...	10.315	11.382	16571451	30280388	18.652m	18.767

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012122\
Data File : PL072158.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2022 12:09
Operator : AR\AJ
Sample : PB142173BL
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PB142173BL


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 01/24/2022
Supervised By : Ankita Jodhani 01/25/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 22 01:57:39 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011722.M
Quant Title : GC Extractables
QLast Update : Tue Jan 18 10:53:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 4.555 min
Delta R.T.: -0.003 min
Response: 9847927
Conc: 17.74 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB142173BL


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/24/2022
Supervised By :Ankita Jodhani 01/25/2022




#1 Tetrachloro-m-xylene

R.T.: 5.491 min
Delta R.T.: -0.003 min
Response: 28882340
Conc: 17.78 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.315 min
Delta R.T.: -0.012 min
Response: 16571451
Conc: 18.65 ng/ml m



#28 Decachlorobiphenyl

R.T.: 11.382 min
Delta R.T.: -0.013 min
Response: 30280388
Conc: 18.77 ng/ml