

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012219\
 Data File : PL044141.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jan 2019 18:50
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 23 02:23:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012219.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 23 00:42:56 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.366	3.985	313.7E6	111.3E6	51.760	51.015
28)	SA Decachlor...	8.048	9.040	310.6E6	102.5E6	44.792	46.225
Target Compounds							
2)	A alpha-BHC	3.797	4.375	472.5E6	144.6E6	52.953	52.102
3)	MA gamma-BHC...	4.074	4.659	422.7E6	133.5E6	52.569	50.501
4)	MA Heptachlor	4.362	5.167	407.6E6	127.8E6	51.701	49.261
5)	MB Aldrin	4.602	5.471	400.8E6	123.6E6	51.331	49.113
6)	B beta-BHC	4.326	4.830	174.8E6	60213780	50.408	51.030
7)	B delta-BHC	4.522	5.046	341.1E6	98542779	52.480	51.840
8)	B Heptachlor...	5.040	5.857	378.9E6	118.1E6	50.186	51.079
9)	A Endosulfan I	5.372	6.213	352.1E6	109.5E6	49.982	47.307
10)	B gamma-Chl...	5.263	6.090	386.9E6	120.5E6	49.950	47.755
11)	B alpha-Chl...	5.320	6.163	379.8E6	119.1E6	49.837	47.329
12)	B 4,4'-DDE	5.487	6.321	347.7E6	104.1E6	51.332	48.525
13)	MA Dieldrin	5.612	6.469	378.6E6	113.6E6	50.380	47.431
14)	MA Endrin	5.865	6.686	323.2E6	91717315	48.204	45.511
15)	B Endosulfa...	6.139	6.895	330.2E6	101.1E6	48.936	47.882
16)	A 4,4'-DDD	5.999	6.810	287.4E6	83702446	51.722	48.809
17)	MA 4,4'-DDT	6.233	7.108	263.2E6	80718389	51.626	49.187
18)	B Endrin al...	6.306	7.019	269.4E6	88668675	47.644	46.484
19)	B Endosulfa...	6.517	7.241	278.8E6	90088055	48.889	46.851
20)	A Methoxychlor	6.778	7.565	143.9E6	49496709	48.513	48.394
21)	B Endrin ke...	7.000	7.711	341.0E6	101.5E6	48.982	47.586

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012219\
Data File : PL044141.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jan 2019 18:50
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 23 02:23:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012219.M
Quant Title : GC Extractables
QLast Update : Wed Jan 23 00:42:56 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

