

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012220\
 Data File : PL055995.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jan 2020 15:02
 Operator : AJ\MA
 Sample : PB126232BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB126232BS

Manual Integrations
 APPROVED

Ankita
 1/23/2020 9:45:59 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 22 16:57:44 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012020.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 02:21:03 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.489	3.966	175.3E6	57249625	18.290	20.256
28)	SA Decachlor...	8.233	8.953	211.2E6	46784575	20.198	20.421
Target Compounds							
2)	A alpha-BHC	3.926	4.349	700.8E6	193.0E6	42.234	45.520
3)	MA gamma-BHC...	4.209	4.629	675.1E6	178.0E6	44.502	44.879
4)	MA Heptachlor	4.511	5.128	676.8E6	160.2E6	45.321	47.374
5)	MB Aldrin	4.759	5.428	682.8E6	160.9E6	46.653	46.933
6)	B beta-BHC	4.459	4.798	290.0E6	79051511	45.193	49.814
7)	B delta-BHC	4.663	5.011	605.4E6	125.8E6	41.302	42.097
8)	B Heptachlo...	5.200	5.809	634.7E6	134.3E6	46.011	44.929
9)	A Endosulfan I	5.539	6.162	594.9E6	133.3E6	45.848	45.793
10)	B gamma-Chl...	5.427	6.041	659.7E6	143.6E6	47.024	46.568
11)	B alpha-Chl...	5.485	6.112	629.3E6	141.0E6	45.524	45.663
12)	B 4,4'-DDE	5.648	6.267	614.4E6	136.4E6	47.039	46.573
13)	MA Dieldrin	5.781	6.416	649.8E6	141.2E6	46.979	46.709
14)	MA Endrin	6.039	6.632	548.7E6	111.4E6	43.433	43.949
15)	B Endosulfa...	6.311	6.839	521.4E6	118.8E6	46.693	44.308
16)	A 4,4'-DDD	6.163	6.752	473.6E6	110.0E6	47.562	47.731
17)	MA 4,4'-DDT	6.403	7.047	487.3E6	104.4E6	48.144	46.252
18)	B Endrin al...	6.478	6.962	455.7E6	99996588	43.849	44.296
19)	B Endosulfa...	6.691	7.183	508.7E6	112.7E6	47.040	45.515
20)	A Methoxychlor	6.944	7.501	238.0E6	53955088	45.600	45.219
21)	B Endrin ke...	7.180	7.649	603.3E6	127.1E6	47.924m	48.125
22)	Mirex	7.371	8.098	371.2E6	83419067	52.856	45.717

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012220\
Data File : PL055995.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jan 2020 15:02
Operator : AJ\MA
Sample : PB126232BS
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB126232BS

Manual Integrations
APPROVED

Ankita
1/23/2020 9:45:59 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 22 16:57:44 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012020.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 02:21:03 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

