

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012221\
 Data File : PL064295.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jan 2021 10:46
 Operator : AR\AJ
 Sample : L5214-04
 Misc : TOX MDL 100 PPB
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 MDL-SOIL-QT1-2021-02

Manual Integrations
 APPROVED

Ankita
 1/22/2021 9:58:28 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 22 04:34:07 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012121CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 21 03:31:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.434	4.108	61878190	47991122	21.279	19.895
27)	SA Decachlor...	8.168	9.271	113.1E6	88207054	37.572	38.004
Target Compounds							
22)	Toxaphene-1	5.676	6.538	3558236	1737865	123.575	106.316
23)	Toxaphene-2	6.273	6.832	8851304	1888995	113.472	105.153m
24)	Toxaphene-3	6.518	7.319	8410982	5251117	99.421	97.678
25)	Toxaphene-4	6.871	7.408	5486289	4795714	88.958	90.795
26)	Toxaphene-5	7.001	7.830	10421773	4239633	123.182	102.744

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012221\
Data File : PL064295.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2021 10:46
Operator : AR\AJ
Sample : L5214-04
Misc : TOX MDL 100 PPB
ALS Vial : 10 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 22 04:34:07 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012121CLP.M
Quant Title : GC Extractables
QLast Update : Thu Jan 21 03:31:19 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Manual Integrations
APPROVED

1/22/2021 9:58:28 AM
Ankita

Instrument :
ECD_L
Client Sample :
MDL-SOIL-QT1-2021-02