

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012320\
 Data File : PL056012.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jan 2020 14:18
 Operator : AJ\MA
 Sample : L1230-03MS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 STOCK-PILE-BMS

Manual Integrations
 APPROVED

Ankita
 1/24/2020 10:42:24 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 23 17:02:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012020.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 02:21:03 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.489	3.966	235.5E6	73604414	24.576	26.043
2)	SA Decachlor...	8.233	8.953	249.2E6	54124325	23.833	23.625
Target Compounds							
2)	A alpha-BHC	3.926	4.349	898.4E6	240.7E6	54.142	56.752
3)	MA gamma-BHC...	4.209	4.629	852.6E6	221.3E6	56.198	55.787
4)	MA Heptachlor	4.510	5.128	845.4E6	196.4E6	56.618	58.084
5)	MB Aldrin	4.759	5.428	820.1E6	192.4E6	56.037	56.135
6)	B beta-BHC	4.459	4.797	334.8E6	95418808	52.166	60.127
7)	B delta-BHC	4.663	5.011	743.6E6	152.6E6	50.733	51.071
8)	B Heptachlo...	5.201	5.809	776.9E6	158.3E6	56.319	52.980
9)	A Endosulfan I	5.539	6.162	698.6E6	161.6E6	53.839	55.489
10)	B gamma-Chl...	5.427	6.041	769.1E6	171.1E6	54.823	55.500
11)	B alpha-Chl...	5.486	6.112	721.3E6	172.0E6	52.177	55.711
12)	B 4,4'-DDE	5.648	6.267	757.7E6	165.1E6	58.013	56.379
13)	MA Dieldrin	5.782	6.416	780.8E6	169.3E6	56.447	56.025
14)	MA Endrin	6.040	6.632	663.4E6	136.3E6	52.512	53.809
15)	B Endosulfa...	6.311	6.839	657.7E6	147.7E6	58.899	55.087
16)	A 4,4'-DDD	6.163	6.753	552.9E6	132.8E6	55.522	57.596
17)	MA 4,4'-DDT	6.403	7.048	591.3E6	128.3E6	58.424	56.847
18)	B Endrin al...	6.479	6.962	558.1E6	125.7E6	53.697	55.673
19)	B Endosulfa...	6.691	7.183	606.9E6	132.0E6	56.117	53.307
20)	A Methoxychlor	6.946	7.502	306.5E6	65527200	58.738	54.917
21)	B Endrin ke...	7.180	7.649	695.7E6	148.3E6	55.262m	56.176
22)	Mirex	7.371	8.099	460.6E6	100.1E6	65.588	54.843

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012320\
Data File : PL056012.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jan 2020 14:18
Operator : AJ\MA
Sample : L1230-03MS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
STOCK-PILE-BMS

Manual Integrations
APPROVED

Ankita
1/24/2020 10:42:24 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 23 17:02:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012020.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 02:21:03 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

