

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012325\
 Data File : PL093771.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jan 2025 12:50
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 23 14:22:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.540	2.776	135.6E6	165.7E6	50.361	50.764
28)	SA Decachlor...	9.056	7.911	106.3E6	175.6E6	50.811	50.109
Target Compounds							
2)	A alpha-BHC	3.995	3.278	192.8E6	255.3E6	50.292	52.216
3)	MA gamma-BHC...	4.328	3.608	186.1E6	248.3E6	50.526	52.362
4)	MA Heptachlor	4.916	3.946	165.1E6	238.1E6	50.377	51.152
5)	MB Aldrin	5.257	4.226	163.6E6	237.8E6	50.001	52.131
6)	B beta-BHC	4.527	3.908	80416942	102.7E6	50.032	51.399
7)	B delta-BHC	4.773	4.136	177.9E6	247.6E6	50.764	52.116
8)	B Heptachlo...	5.684	4.728	145.9E6	216.9E6	49.071	51.895
9)	A Endosulfan I	6.070	5.098	129.7E6	198.5E6	49.063	51.211
10)	B gamma-Chl...	5.940	4.978	139.2E6	220.0E6	49.957	51.910
11)	B alpha-Chl...	6.019	5.042	138.3E6	217.2E6	49.614	51.881
12)	B 4,4'-DDE	6.193	5.231	126.8E6	213.9E6	52.082	53.347
13)	MA Dieldrin	6.344	5.362	137.6E6	222.5E6	49.579	51.808
14)	MA Endrin	6.575	5.637	114.7E6	183.8E6	48.913	49.778
15)	B Endosulfa...	6.795	5.932	118.7E6	190.8E6	49.268	51.525
16)	A 4,4'-DDD	6.710	5.785	102.5E6	174.9E6	53.911	55.407
17)	MA 4,4'-DDT	7.024	6.035	103.1E6	172.1E6	52.297	52.892
18)	B Endrin al...	6.925	6.111	98485359	154.1E6	50.660	50.624
19)	B Endosulfa...	7.159	6.334	113.1E6	181.6E6	49.950	50.938
20)	A Methoxychlor	7.500	6.610	56467361	88752734	54.119	49.634
21)	B Endrin ke...	7.644	6.839	128.8E6	218.4E6	51.054	52.056
22)	Mirex	8.117	7.019	103.4E6	172.8E6	49.649	51.093

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012325\
Data File : PL093771.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jan 2025 12:50
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 23 14:22:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

