

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012424\
 Data File : PL087930.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jan 2024 14:37
 Operator : AR\AJ
 Sample : P1187-02
 Misc : CHLOR LOQ 50PPB
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 LOQ-SOIL-02-QT1-2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 25 01:01:36 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122823.M
 Quant Title : GC Extractables
 QLast Update : Fri Dec 29 03:17:47 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.422	2.673	61481530	27470974	20.537	22.678
28)	SA Decachlor...	8.981	7.859	65881260	30422561	23.132	24.015
Target Compounds							
23)	Chlordane-1	4.588	3.673	6305104	2147614	47.238	45.950
24)	Chlordane-2	5.123	4.256	7969233	2862087	54.040	51.156
25)	Chlordane-3	5.837	4.890	26498120	7370372	47.636	46.572
26)	Chlordane-4	5.921	4.953	33386511	7774097	48.266	45.500
27)	Chlordane-5	6.770	5.626	5620255	2269566	49.837m	48.215

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012424\
Data File : PL087930.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jan 2024 14:37
Operator : AR\AJ
Sample : P1187-02
Misc : CHLOR LOQ 50PPB
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
LOQ-SOIL-02-QT1-2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 25 01:01:36 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL122823.M
Quant Title : GC Extractables
QLast Update : Fri Dec 29 03:17:47 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

