

Method Path : P:\HPCHEM1\ECD_L\methods\
 Method File : LTX012618.M
 Title : GC Extractables
 Last Update : Sat Jan 27 04:53:55 2018
 Response Via : Initial Calibration

Calibration Files

500 =PL032509.D 1000 =PL032507.D 250 =PL032510.D
 750 =PL032508.D =

Compound			500	1000	250	750	Avg		%RSD
1) SA	Tetrachloro-m-xylene		8.178	8.138	8.026	8.351	8.173	E6	1.65
2)	Toxaphene-1		7.211	7.168	6.759	7.225	7.091	E4	3.14
3)	Toxaphene-2		1.050	1.338	1.412	1.595	1.349	E5	16.82
4)	Toxaphene-3		6.308	6.219	6.045	6.640	6.303	E4	3.96
5)	Toxaphene-4		1.489	1.530	1.463	1.553	1.509	E5	2.67
6)	Toxaphene-5		1.430	1.722	1.721	1.904	1.694	E5	11.56
7) SA	Decachlorobiphenyl		1.049	1.013	1.085	1.067	1.053	E7	2.90

Signal #2 Calibration Files

500 =PL032509.D 1000 =PL032507.D 250 =PL032510.D
 750 =PL032508.D =

Compound		500	1000	250	750	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.164	1.128	1.166	1.171	1.158	E6	1.72
2)	Toxaphene-1	1.914	1.853	1.914	1.937	1.905	E4	1.89
3)	Toxaphene-2	1.204	1.179	1.261	1.215	1.215	E4	2.80
4)	Toxaphene-3	1.400	1.383	1.396	1.444	1.406	E4	1.90
5)	Toxaphene-4	3.796	3.827	3.578	3.938	3.785	E4	3.98
6)	Toxaphene-5	1.834	1.805	1.918	1.883	1.860	E4	2.69
7) SA	Decachlorobiphenyl	1.737	1.645	1.832	1.745	1.740	E6	4.40

(#) = Out of Range