

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012622\
 Data File : PL072266.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jan 2022 09:39
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 27 00:17:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011722.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 18 10:53:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.556	5.492	26362075	78801737	47.483	48.520
28)	SA Decachlor...	10.317	11.383	40815680	78094298	45.941	48.400
Target Compounds							
2)	A alpha-BHC	5.262	6.073	37069859	115.2E6	47.942	47.738
3)	MA gamma-BHC...	5.685	6.468	36348126	107.7E6	48.206	47.858m
4)	MA Heptachlor	6.091	7.115	41612466	108.6E6	48.235	46.850
5)	MB Aldrin	6.415	7.485	38145151	103.2E6	48.307	46.674
6)	B beta-BHC	6.071	6.729	15736765	47281280	47.606	47.753
7)	B delta-BHC	6.328	6.995	38553422	108.4E6	48.670	44.044
8)	B Heptachlo...	6.993	7.953	36950224	94047852	48.540	46.970
9)	A Endosulfan I	7.394	8.358	34518434	84566777	48.569	46.720
10)	B gamma-Chl...	7.269	8.221	37707798	93216577	47.991	46.128m
11)	B alpha-Chl...	7.339	8.303	37632968	91855520	47.971	46.326
12)	B 4,4'-DDE	7.554	8.493	35759283	87927439	49.529	46.492
13)	MA Dieldrin	7.681	8.646	37118014	89754504	48.993	45.686
14)	MA Endrin	7.972	8.885	33071318	75403101	47.200	45.478
15)	B Endosulfa...	8.284	9.120	34821784	78245185	46.833	45.626m
16)	A 4,4'-DDD	8.140	9.033	29407025	68853278	50.164	48.156
17)	MA 4,4'-DDT	8.400	9.350	31628844	72893707	47.629	46.119
18)	B Endrin al...	8.473	9.256	26886907	59379548	47.191	47.834
19)	B Endosulfa...	8.704	9.494	34732944	73440352	47.373	47.042
20)	A Methoxychlor	8.996	9.837	19773898	38617204	46.869	46.620
21)	B Endrin ke...	9.222	9.991	40888853	80410960	49.104	48.587
22)	Mirex	9.404	10.454	34656018	57496058	46.929	46.079

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012622\
 Data File : PL072266.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jan 2022 09:39
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 27 00:17:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011722.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 18 10:53:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

