

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012722\
 Data File : PL072273.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jan 2022 09:23
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 01/28/2022
 Supervised By :Ankita Jodhani 01/28/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 28 00:04:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011722.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 18 10:53:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.555	5.492	26266479	78670400	47.311	48.439
28) SA Decachlor...	10.316	11.383	40364315	75139564	45.433	46.569
Target Compounds						
2) A alpha-BHC	5.262	6.073	37174738	115.4E6	48.078	47.847
3) MA gamma-BHC...	5.685	6.468	36240595	107.3E6	48.063	47.695m
4) MA Heptachlor	6.091	7.115	41483387	107.6E6	48.086	46.396
5) MB Aldrin	6.414	7.485	37668801	102.1E6	47.704	46.151
6) B beta-BHC	6.070	6.729	15334315	46881106	46.388	47.349
7) B delta-BHC	6.328	6.994	38048068	107.4E6	48.032	43.635
8) B Heptachlo...	6.992	7.953	36485925	93259501	47.930	46.576
9) A Endosulfan I	7.394	8.357	34200638	83638571	48.122	46.207
10) B gamma-Chl...	7.269	8.222	37320270	93666670	47.498	46.351
11) B alpha-Chl...	7.339	8.302	37353085	91242139	47.614	46.017
12) B 4,4'-DDE	7.554	8.493	35190068	87128250	48.741	46.069
13) MA Dieldrin	7.680	8.646	36670656	89284445	48.403	45.447
14) MA Endrin	7.972	8.885	32586895	73824338	46.509	44.526
15) B Endosulfa...	8.284	9.121	34436713	77714190	46.316	45.317
16) A 4,4'-DDD	8.140	9.033	28717599	67171952	48.988	46.980
17) MA 4,4'-DDT	8.400	9.350	31213412	72311801	47.003	45.751
18) B Endrin al...	8.473	9.256	26635090	58563138	46.749	47.176
19) B Endosulfa...	8.703	9.494	34385773	72483314	46.899	46.429
20) A Methoxychlor	8.996	9.837	19313847	37865051	45.779	45.712
21) B Endrin ke...	9.221	9.990	40251394	79824580	48.339	48.232
22) Mirex	9.403	10.453	34386611	57044473	46.564	45.717

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012722\
 Data File : PL072273.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jan 2022 09:23
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 01/28/2022
 Supervised By :Ankita Jodhani 01/28/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 28 00:04:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011722.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 18 10:53:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

