

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012723\
 Data File : PL080584.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jan 2023 22:34
 Operator : AR\AJ
 Sample : 01270-02MS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 JC-1-012323MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 01/30/2023
 Supervised By : Ankita Jodhani 01/30/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 28 00:14:07 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012023.M
 Quant Title : GC Extractables
 QLast Update : Sun Jan 22 23:56:48 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.352	2.689	36120296	35949510	16.841	16.367
28) SA Decachlor...	8.787	7.836	29780952	37016202	17.010	15.536
Target Compounds						
2) A alpha-BHC	3.800	3.190	151.7E6	145.8E6	47.869	45.840
3) MA gamma-BHC...	4.128	3.518	139.8E6	135.4E6	47.572	45.338
4) MA Heptachlor	4.712	3.859	134.9E6	135.7E6	45.640	42.656
5) MB Aldrin	5.050	4.138	129.7E6	132.4E6	47.426	44.894
6) B beta-BHC	4.324	3.814	60505948	58389495	44.326	43.750
7) B delta-BHC	4.567	4.043	140.3E6	133.1E6	48.453	46.131
8) B Heptachlo...	5.474	4.641	120.0E6	122.1E6	46.390	43.870
9) A Endosulfan I	5.856	5.010	108.0E6	111.0E6	45.746	42.671
10) B gamma-Chl...	5.728	4.890	118.8E6	124.4E6	47.034	43.909m
11) B alpha-Chl...	5.808	4.956	117.1E6	121.4E6	46.389	42.384
12) B 4,4'-DDE	5.986	5.148	103.9E6	104.8E6	46.636	44.450
13) MA Dieldrin	6.130	5.275	115.1E6	120.6E6	48.096	44.584
14) MA Endrin	6.357	5.549	98089727	104.6E6	45.355	45.050
15) B Endosulfa...	6.576	5.843	98394827	105.9E6	42.924	50.827
16) A 4,4'-DDD	6.500	5.702	87846010	90577701	47.951	48.200
17) MA 4,4'-DDT	6.812	5.953	83790050	86593142	40.329	41.206
18) B Endrin al...	6.705	6.023	82488079	87107612	49.590	47.701
19) B Endosulfa...	6.939	6.246	97593472	104.3E6	47.149	44.868
20) A Methoxychlor	7.290	6.531	46555611	49843470	42.197	41.075
21) B Endrin ke...	7.419	6.751	106.7E6	118.7E6	48.369	46.481
22) Mirex	7.891	6.934	82023842	100.1E6	47.694	43.978

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012723\
Data File : PL080584.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jan 2023 22:34
Operator : AR\AJ
Sample : 01270-02MS
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

JC-1-012323MS

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 01/30/2023

Supervised By :Ankita Jodhani 01/30/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 28 00:14:07 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012023.M
Quant Title : GC Extractables
QLast Update : Sun Jan 22 23:56:48 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

