

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012819\
 Data File : PL044287.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jan 2019 20:20
 Operator : AJ\SJ
 Sample : K1261-04
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 MH-C05

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 00:04:14 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012219.M
 Quant Title : GC Extractables
 QLast Update : Mon Jan 28 01:20:09 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.367	3.986	141.8E6	50830806	23.399	23.303
28) SA Decachlor...	8.052	9.047	155.5E6	47097687	22.418	21.231
Target Compounds						
6) B beta-BHC	4.286f	0.000	16169804	0	4.664	N.D. #
15) B Endosulfa...	6.093f	0.000	15653852	0	2.320	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012819\
Data File : PL044287.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jan 2019 20:20
Operator : AJ\SJ
Sample : K1261-04
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
MH-C05

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 29 00:04:14 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012219.M
Quant Title : GC Extractables
QLast Update : Mon Jan 28 01:20:09 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.367 min
Delta R.T.: 0.002 min
Response: 141808549
Conc: 23.40 ng/ml

Instrument :
ECD_L
ClientSampleId :
MH-C05




#1 Tetrachloro-m-xylene

R.T.: 3.986 min
Delta R.T.: 0.002 min
Response: 50830806
Conc: 23.30 ng/ml




#6 beta-BHC

R.T.: 4.286 min
Delta R.T.: -0.039 min
Response: 16169804
Conc: 4.66 ng/ml



#6 beta-BHC


R.T.: 0.000 min
Exp R.T. : 4.830 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 6.093 min
Delta R.T.: -0.045 min
Response: 15653852
Conc: 2.32 ng/ml

Instrument :
ECD_L
ClientSampleId :
MH-C05




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 6.895 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.052 min
Delta R.T.: 0.004 min
Response: 155469736
Conc: 22.42 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.047 min
Delta R.T.: 0.008 min
Response: 47097687
Conc: 21.23 ng/ml