

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012820\
 Data File : PL056159.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jan 2020 22:04
 Operator : AJ\MA
 Sample : L1242-08
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 03:16:17 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012820.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 28 13:10:37 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.499	4.016	208.3E6	57627116	19.893	19.708
28)	SA Decachlor...	8.255	9.094	123.9E6	31269859	11.772	13.824
Target Compounds							
10)	B gamma-Chl...	5.462f	6.135	11533798	8022853	0.735	2.523 #
25)	Chlordane-3	5.462f	6.135	11533798	8022853	9.135	21.347 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012820\
Data File : PL056159.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jan 2020 22:04
Operator : AJ\MA
Sample : L1242-08
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 29 03:16:17 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012820.M
Quant Title : GC Extractables
QLast Update : Tue Jan 28 13:10:37 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.499 min
Delta R.T.: 0.002 min
Response: 208343479
Conc: 19.89 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.016 min
Delta R.T.: 0.000 min
Response: 57627116
Conc: 19.71 ng/ml




#10 gamma-Chlordane

R.T.: 5.462 min
Delta R.T.: 0.021 min
Response: 11533798
Conc: 0.74 ng/ml



#10 gamma-Chlordane


R.T.: 6.135 min
Delta R.T.: 0.011 min
Response: 8022853
Conc: 2.52 ng/ml



#25 Chlordane-3


R.T.: 5.462 min
Delta R.T.: 0.017 min
Response: 11533798
Conc: 9.14 ng/ml

Instrument :
ECD_L
ClientSampleId :




#25 Chlordane-3

R.T.: 6.135 min
Delta R.T.: 0.010 min
Response: 8022853
Conc: 21.35 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.255 min
Delta R.T.: 0.002 min
Response: 123934161
Conc: 11.77 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.094 min
Delta R.T.: -0.002 min
Response: 31269859
Conc: 13.82 ng/ml