

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012822\
 Data File : PL072311.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jan 2022 14:01
 Operator : AR\AJ
 Sample : N1309-11
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 STATE-ST-SP

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 01/31/2022
 Supervised By : Ankita Jodhani 01/31/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 01:26:52 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011722.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 18 10:53:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.557	5.493	6108378	18051323	11.002m	11.115
2)	SA Decachlor...	10.319	11.387	10191839	16931567	11.472	10.494

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012822\
Data File : PL072311.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jan 2022 14:01
Operator : AR\AJ
Sample : N1309-11
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
STATE-ST-SP


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 01/31/2022
Supervised By : Ankita Jodhani 01/31/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 29 01:26:52 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL011722.M
Quant Title : GC Extractables
QLast Update : Tue Jan 18 10:53:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 4.557 min
Delta R.T.: -0.001 min
Response: 6108378
Conc: 11.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
STATE-ST-SP


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 01/31/2022
Supervised By :Ankita Jodhani 01/31/2022




#1 Tetrachloro-m-xylene

R.T.: 5.493 min
Delta R.T.: -0.001 min
Response: 18051323
Conc: 11.11 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.319 min
Delta R.T.: -0.007 min
Response: 10191839
Conc: 11.47 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.387 min
Delta R.T.: -0.008 min
Response: 16931567
Conc: 10.49 ng/ml