

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX012921.M
 Title : GC Extractables
 Last Update : Fri Jan 29 05:16:43 2021
 Response Via : Initial Calibration

Calibration Files

500 =PL064498.D 1000 =PL064496.D 250 =PL064499.D
 750 =PL064497.D 100 =PL064500.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	3.009	3.029	2.963	3.024	2.974	3.000	E6	0.99
2)	Toxaphene-1	2.272	2.312	2.421	2.271	2.576	2.370	E4	5.50
3)	Toxaphene-2	2.125	1.964	2.039	1.921	1.775	1.965	E4	6.69
4)	Toxaphene-3	5.757	5.852	6.566	5.952	5.982	6.022	E4	5.27
5)	Toxaphene-4	4.081	4.260	4.017	4.304	3.766	4.086	E4	5.27
6)	Toxaphene-5	6.340	6.357	7.129	6.511	6.321	6.532	E4	5.24
7) SA	Decachlorobiphenyl	2.974	2.869	3.160	2.907	3.377	3.057	E6	6.89

Signal #2 Calibration Files

500 =PL064498.D 1000 =PL064496.D 250 =PL064499.D
 750 =PL064497.D 100 =PL064500.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.396	2.447	2.331	2.399	2.274	2.369	E6	2.84
2)	Toxaphene-1	1.926	1.910	1.897	1.895	1.773	1.881	E4	3.25
3)	Toxaphene-2	2.072	2.039	2.040	2.029	1.848	2.006	E4	4.48
4)	Toxaphene-3	3.746	3.701	3.731	3.677	3.276	3.626	E4	5.45
5)	Toxaphene-4	3.527	3.579	3.517	3.565	3.060	3.450	E4	6.36
6)	Toxaphene-5	2.991	2.900	3.061	2.931	2.636	2.904	E4	5.57
7) SA	Decachlorobiphenyl	2.009	1.915	2.095	1.965	2.167	2.030	E6	4.99

(#) = Out of Range