

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012925\
 Data File : PL093843.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jan 2025 10:49
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 30 00:22:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.539	2.775	133.0E6	159.1E6	49.403	48.735
28) SA Decachlor...	9.061	7.914	108.5E6	182.4E6	51.875	52.043
Target Compounds						
2) A alpha-BHC	3.995	3.277	192.7E6	246.5E6	50.270	50.417
3) MA gamma-BHC...	4.328	3.608	182.8E6	232.9E6	49.632	49.119
4) MA Heptachlor	4.916	3.946	168.6E6	235.3E6	51.448	50.553
5) MB Aldrin	5.258	4.226	160.7E6	220.6E6	49.109	48.359
6) B beta-BHC	4.526	3.908	81410522	100.7E6	50.650	50.423
7) B delta-BHC	4.774	4.137	173.7E6	231.4E6	49.542	48.696
8) B Heptachlo...	5.685	4.729	144.7E6	204.8E6	48.668	48.983
9) A Endosulfan I	6.071	5.099	132.4E6	191.7E6	50.096	49.444
10) B gamma-Chl...	5.942	4.979	142.7E6	218.7E6	51.185	51.609
11) B alpha-Chl...	6.020	5.043	142.3E6	214.1E6	51.024	51.150
12) B 4,4'-DDE	6.194	5.231	130.6E6	210.6E6	53.648	52.526
13) MA Dieldrin	6.347	5.363	140.1E6	217.3E6	50.480	50.587
14) MA Endrin	6.576	5.639	123.2E6	190.6E6	52.539	51.609
15) B Endosulfa...	6.796	5.934	121.6E6	191.9E6	50.479	51.807
16) A 4,4'-DDD	6.712	5.787	102.8E6	169.4E6	54.070	53.661
17) MA 4,4'-DDT	7.026	6.037	111.2E6	186.3E6	56.411	57.251
18) B Endrin al...	6.926	6.113	96187663	147.5E6	49.478	48.460
19) B Endosulfa...	7.161	6.337	114.3E6	181.8E6	50.493	50.978
20) A Methoxychlor	7.503	6.612	59300697	98340551	56.834	54.996
21) B Endrin ke...	7.647	6.842	127.7E6	212.8E6	50.606	50.723
22) Mirex	8.120	7.022	101.3E6	165.6E6	48.625	48.955

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012925\
Data File : PL093843.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jan 2025 10:49
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 30 00:22:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

