

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012925\
 Data File : PL093862.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jan 2025 18:45
 Operator : AR\AJ
 Sample : Q1206-04MSD
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 JPP-20.1-012725MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 30 00:29:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 21 14:58:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.539	2.775	59976596	69783961	22.273	21.379
2) SA Decachlor...	9.058	7.912	65138738	107.6E6	31.138	30.695
Target Compounds						
2) A alpha-BHC	3.995	3.277	221.8E6	291.0E6	57.849	59.515
3) MA gamma-BHC...	4.327	3.607	214.6E6	278.4E6	58.271	58.730
4) MA Heptachlor	4.916	3.946	198.3E6	276.4E6	60.509	59.386
5) MB Aldrin	5.257	4.225	187.1E6	257.8E6	57.170	56.511
6) B beta-BHC	4.526	3.908	96284344	120.3E6	59.904	60.212
7) B delta-BHC	4.773	4.136	214.7E6	283.7E6	61.262	59.718
8) B Heptachlo...	5.685	4.728	177.5E6	254.2E6	59.678	60.805
9) A Endosulfan I	6.070	5.098	162.1E6	243.5E6	61.318	62.820
10) B gamma-Chl...	5.940	4.978	174.7E6	268.9E6	62.665	63.453
11) B alpha-Chl...	6.019	5.042	174.6E6	263.4E6	62.600	62.919
12) B 4,4'-DDE	6.193	5.231	159.1E6	256.9E6	65.351	64.063
13) MA Dieldrin	6.345	5.362	173.7E6	270.6E6	62.591	62.994
14) MA Endrin	6.576	5.638	150.5E6	245.3E6	64.201	66.422
15) B Endosulfa...	6.795	5.933	150.9E6	240.6E6	62.620	64.961
16) A 4,4'-DDD	6.711	5.786	126.7E6	210.0E6	66.658	66.517
17) MA 4,4'-DDT	7.025	6.036	139.9E6	231.5E6	70.964	71.150
18) B Endrin al...	6.925	6.112	117.6E6	185.2E6	60.488	60.818
19) B Endosulfa...	7.160	6.335	141.4E6	231.7E6	62.475	64.966
20) A Methoxychlor	7.502	6.611	72191075	120.0E6	69.189	67.122
21) B Endrin ke...	7.645	6.840	159.2E6	270.4E6	63.091	64.463
22) Mirex	8.119	7.020	122.8E6	201.6E6	58.987	59.614

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL012925\
Data File : PL093862.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jan 2025 18:45
Operator : AR\AJ
Sample : Q1206-04MSD
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
JPP-20.1-012725MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 30 00:29:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012125.M
Quant Title : GC Extractables
QLast Update : Tue Jan 21 14:58:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

