

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013019\
 Data File : PL044347.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jan 2019 21:24
 Operator : AJ\SJ
 Sample : PB116723BL
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 08:34:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012219.M
 Quant Title : GC Extractables
 QLast Update : Mon Jan 28 01:20:09 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.367	3.986	148.9E6	52097864	24.574	23.884
28)	SA Decachlor...	8.052	9.047	154.1E6	48322583	22.217	21.783
Target Compounds							
6)	B beta-BHC	4.285f	0.000	6723573	0	1.939	N.D. #
10)	B gamma-Chl...	0.000	6.117	0	15763337	N.D.	6.247 #
11)	B alpha-Chl...	0.000	6.117f	0	15763337	N.D.	6.265 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013019\
Data File : PL044347.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2019 21:24
Operator : AJ\SJ
Sample : PB116723BL
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 08:34:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012219.M
Quant Title : GC Extractables
QLast Update : Mon Jan 28 01:20:09 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.367 min
Delta R.T.: -0.005 min
Response: 148927573
Conc: 24.57 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.986 min
Delta R.T.: -0.001 min
Response: 52097864
Conc: 23.88 ng/ml




#6 beta-BHC

R.T.: 4.285 min
Delta R.T.: -0.049 min
Response: 6723573
Conc: 1.94 ng/ml



#6 beta-BHC


R.T.: 0.000 min
Exp R.T. : 4.835 min
Response: 0
Conc: N.D.



#10 gamma-Chlordane


R.T.: 0.000 min
Exp R.T.: 5.273 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#10 gamma-Chlordane

R.T.: 6.117 min
Delta R.T.: 0.020 min
Response: 15763337
Conc: 6.25 ng/ml




#11 alpha-Chlordane

R.T.: 0.000 min
Exp R.T.: 5.329 min
Response: 0
Conc: N.D.



#11 alpha-Chlordane


R.T.: 6.117 min
Delta R.T.: -0.053 min
Response: 15763337
Conc: 6.27 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.052 min
Delta R.T.: -0.009 min
Response: 154076154
Conc: 22.22 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 9.047 min
Delta R.T.: -0.005 min
Response: 48322583
Conc: 21.78 ng/ml