

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013021\
 Data File : PL064507.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jan 2021 09:43
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 30 03:19:31 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012921.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 29 04:48:57 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.433	4.107	166.0E6	128.3E6	54.526	53.881
28)	SA Decachlor...	8.163	9.268	160.1E6	105.7E6	50.807	51.645
Target Compounds							
2)	A alpha-BHC	3.866	4.509	283.5E6	192.4E6	56.511	55.659
3)	MA gamma-BHC...	4.146	4.800	261.2E6	180.2E6	55.780	54.645
4)	MA Heptachlor	4.445	5.313	251.3E6	183.9E6	54.350	52.752
5)	MB Aldrin	4.691	5.624	231.0E6	159.7E6	54.370	53.510
6)	B beta-BHC	4.395	4.977	104.4E6	88825011	54.000	52.491
7)	B delta-BHC	4.598	5.198	257.0E6	166.1E6	55.230	53.924
8)	B Heptachlo...	5.131	6.017	218.6E6	138.2E6	53.671	53.961
9)	A Endosulfan I	5.467	6.380	200.5E6	139.2E6	53.105	52.340
10)	B gamma-Chl...	5.357	6.254	219.5E6	153.7E6	53.010	52.865
11)	B alpha-Chl...	5.416	6.326	215.7E6	154.2E6	52.602	52.396
12)	B 4,4'-DDE	5.582	6.483	203.9E6	137.3E6	54.083	53.035
13)	MA Dieldrin	5.709	6.641	215.3E6	144.9E6	53.836	52.557
14)	MA Endrin	5.964	6.863	189.3E6	112.7E6	51.745	49.076
15)	B Endosulfa...	6.237	7.076	191.1E6	127.8E6	52.839	51.783
16)	A 4,4'-DDD	6.095	6.980	173.5E6	119.6E6	55.132	51.967
17)	MA 4,4'-DDT	6.335	7.281	161.2E6	106.0E6	52.535	51.626
18)	B Endrin al...	6.404	7.201	164.1E6	109.0E6	53.505	51.342
19)	B Endosulfa...	6.617	7.425	191.5E6	123.5E6	52.691	51.369
20)	A Methoxychlor	6.881	7.740	90903239	63992370	50.325	50.581
21)	B Endrin ke...	7.103	7.904	210.0E6	127.2E6	53.881	52.317
22)	Mirex	7.290	8.357	152.2E6	98093264	50.790	51.077

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013021\
Data File : PL064507.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jan 2021 09:43
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 30 03:19:31 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012921.M
Quant Title : GC Extractables
QLast Update : Fri Jan 29 04:48:57 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

