

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013024\
 Data File : PL088045.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jan 2024 13:06
 Operator : AR\AJ
 Sample : PCHLORCCC500
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PCHLORCCC500

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 01/31/2024
 Supervised By :Ankita Jodhani 01/31/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 30 21:18:29 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012724.M
 Quant Title : GC Extractables
 QLast Update : Mon Jan 29 12:34:13 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.424	2.673	130.7E6	68449061	45.065	55.146
28)	SA Decachlor...	8.982	7.858	116.9E6	55343210	43.616	45.505
Target Compounds							
23)	Chlordane-1	4.589	3.673	56566720	20338512	432.835	440.691
24)	Chlordane-2	5.124	4.255	62093843	27722909	434.289	461.285
25)	Chlordane-3	5.838	4.889	240.7E6	73622778	450.497	465.971
26)	Chlordane-4	5.922	4.952	293.6E6	76118435	446.365	455.682
27)	Chlordane-5	6.772	5.624	45350943	20576121	413.698m	451.111m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013024\
Data File : PL088045.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2024 13:06
Operator : AR\AJ
Sample : PCHLORCCC500
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 01/31/2024
Supervised By : Ankita Jodhani 01/31/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 30 21:18:29 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012724.M
Quant Title : GC Extractables
QLast Update : Mon Jan 29 12:34:13 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

