

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013123\
 Data File : PL080622.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jan 2023 14:03
 Operator : AR\AJ
 Sample : 01328-02MS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 COMP-1-BORING-1-0-7MS

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 02/01/2023
 Supervised By :Ankita Jodhani 02/01/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 31 14:47:44 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012023.M
 Quant Title : GC Extractables
 QLast Update : Sun Jan 22 23:56:48 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.352	2.689	36041629	37032285	16.804	16.860
28)	SA Decachlor...	8.788	7.835	33793211	43053079	19.302	18.070
Target Compounds							
2)	A alpha-BHC	3.800	3.190	152.7E6	157.1E6	48.198	49.396
3)	MA gamma-BHC...	4.127	3.518	145.6E6	146.4E6	49.538m	49.007
4)	MA Heptachlor	4.712	3.859	140.8E6	148.0E6	47.651	46.499
5)	MB Aldrin	5.050	4.138	132.1E6	142.7E6	48.286	48.364
6)	B beta-BHC	4.324	3.814	61894823	61665144	45.343	46.204
7)	B delta-BHC	4.567	4.043	146.3E6	143.2E6	50.520	49.639
8)	B Heptachlo...	5.475	4.640	125.8E6	131.9E6	48.608	47.402
9)	A Endosulfan I	5.856	5.009	111.6E6	119.8E6	47.270	46.064
10)	B gamma-Chl...	5.730	4.890	121.5E6	134.3E6	48.104	47.392m
11)	B alpha-Chl...	5.808	4.955	120.6E6	130.8E6	47.759	45.642
12)	B 4,4'-DDE	5.986	5.147	106.8E6	113.6E6	47.951	48.174m
13)	MA Dieldrin	6.130	5.274	119.0E6	131.6E6	49.708	48.646m
14)	MA Endrin	6.358	5.549	106.2E6	117.7E6	49.117	50.696
15)	B Endosulfa...	6.577	5.842	101.7E6	113.5E6	44.360	54.467m
16)	A 4,4'-DDD	6.500	5.702	87104010	90719548	47.546	48.276
17)	MA 4,4'-DDT	6.813	5.952	92878889	99457671	44.703	47.327m
18)	B Endrin al...	6.706	6.023	84820313	92177331	50.992	50.477
19)	B Endosulfa...	6.940	6.247	101.0E6	110.9E6	48.798	47.670
20)	A Methoxychlor	7.292	6.531	51524705	54913847	46.701	45.254
21)	B Endrin ke...	7.420	6.751	106.7E6	122.7E6	48.378	48.051
22)	Mirex	7.891	6.935	84696091	107.9E6	49.248	47.401

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL013123\
Data File : PL080622.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jan 2023 14:03
Operator : AR\AJ
Sample : 01328-02MS
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
COMP-1-BORING-1-0-7MS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 02/01/2023
Supervised By :Ankita Jodhani 02/01/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 31 14:47:44 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012023.M
Quant Title : GC Extractables
QLast Update : Sun Jan 22 23:56:48 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

