

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL020119\
 Data File : PL044375.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Feb 2019 15:57
 Operator : AJ\SJ
 Sample : K1297-06
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 A0C4-SS1

Manual Integrations
 APPROVED

Sohil
 2/4/2019 10:14:44 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 23:10:43 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012219.M
 Quant Title : GC Extractables
 QLast Update : Mon Jan 28 01:20:09 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.366	3.985	80963253	29130549	13.360	13.355
28)	SA Decachlor...	8.049	9.044	118.4E6	35649921	17.074	16.070
Target Compounds							
12)	B 4,4'-DDE	5.485	6.321	4588088	2466707	0.677m	1.149 #
17)	MA 4,4'-DDT	6.231	7.108	9617805	1983046	1.887	1.208m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL020119\
Data File : PL044375.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Feb 2019 15:57
Operator : AJ\SJ
Sample : K1297-06
Misc :
ALS Vial : 14 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
A0C4-SS1


Manual Integrations
APPROVED

Sohil
2/4/2019 10:14:44 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 23:10:43 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012219.M
Quant Title : GC Extractables
QLast Update : Mon Jan 28 01:20:09 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.366 min
Delta R.T.: 0.001 min
Response: 80963253
Conc: 13.36 ng/ml

Instrument :
ECD_L
ClientSampleId :
A0C4-SS1


Manual Integrations
APPROVED

Sohil
2/4/2019 10:14:44 AM




#1 Tetrachloro-m-xylene

R.T.: 3.985 min
Delta R.T.: 0.000 min
Response: 29130549
Conc: 13.35 ng/ml




#12 4,4'-DDE

R.T.: 5.485 min
Delta R.T.: 0.000 min
Response: 4588088
Conc: 0.68 ng/ml m



#12 4,4'-DDE

R.T.: 6.321 min
Delta R.T.: 0.002 min
Response: 2466707
Conc: 1.15 ng/ml




#17 4,4'-DDT

R.T.: 6.231 min
Delta R.T.: -0.001 min
Response: 9617805
Conc: 1.89 ng/ml

Instrument :
ECD_L
ClientSampleId :
A0C4-SS1


Manual Integrations
APPROVED

Sohil
2/4/2019 10:14:44 AM




#17 4,4'-DDT

R.T.: 7.108 min
Delta R.T.: 0.001 min
Response: 1983046
Conc: 1.21 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.049 min
Delta R.T.: 0.002 min
Response: 118406118
Conc: 17.07 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.044 min
Delta R.T.: 0.004 min
Response: 35649921
Conc: 16.07 ng/ml