

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL020419\
 Data File : PL044421.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Feb 2019 18:55
 Operator : AJ\SJ
 Sample : PB116798TB
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB116798TB

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 05 00:10:49 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012219.M
 Quant Title : GC Extractables
 QLast Update : Mon Jan 28 01:20:09 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.367	3.986	209.2E6	68821610	34.517	31.551
28)	SA Decachlor...	8.050	9.045	226.0E6	68010787	32.592	30.658
Target Compounds							
6)	B beta-BHC	4.285f	0.000	6452386	0	1.861	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL020419\
Data File : PL044421.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Feb 2019 18:55
Operator : AJ\SJ
Sample : PB116798TB
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB116798TB

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 05 00:10:49 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012219.M
Quant Title : GC Extractables
QLast Update : Mon Jan 28 01:20:09 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.367 min
Delta R.T.: 0.002 min
Response: 209184519
Conc: 34.52 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB116798TB




#1 Tetrachloro-m-xylene

R.T.: 3.986 min
Delta R.T.: 0.001 min
Response: 68821610
Conc: 31.55 ng/ml




#6 beta-BHC

R.T.: 4.285 min
Delta R.T.: -0.040 min
Response: 6452386
Conc: 1.86 ng/ml



#6 beta-BHC


R.T.: 0.000 min
Exp R.T. : 4.830 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.050 min
Delta R.T.: 0.002 min
Response: 226025216
Conc: 32.59 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB116798TB



#28 Decachlorobiphenyl

R.T.: 9.045 min
Delta R.T.: 0.006 min
Response: 68010787
Conc: 30.66 ng/ml