

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL020619\
 Data File : PL044487.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Feb 2019 18:59
 Operator : AJ\SJ
 Sample : K1346-04
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 07 02:56:37 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020619.M
 Quant Title : GC Extractables
 QLast Update : Thu Feb 07 01:39:50 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.367	3.986	124.2E6	41727391	17.447	17.762
28)	SA Decachlor...	8.051	9.045	119.2E6	37501070	16.340	16.932
Target Compounds							
6)	B beta-BHC	4.284f	0.000	6162910	0	1.458	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL020619\
Data File : PL044487.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Feb 2019 18:59
Operator : AJ\SJ
Sample : K1346-04
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 07 02:56:37 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020619.M
Quant Title : GC Extractables
QLast Update : Thu Feb 07 01:39:50 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.367 min
Delta R.T.: 0.000 min
Response: 124153488
Conc: 17.45 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.986 min
Delta R.T.: 0.001 min
Response: 41727391
Conc: 17.76 ng/ml




#6 beta-BHC

R.T.: 4.284 min
Delta R.T.: -0.041 min
Response: 6162910
Conc: 1.46 ng/ml



#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.831 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl


R.T.: 8.051 min

Delta R.T.: 0.002 min

Response: 119176607

Conc: 16.34 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 9.045 min

Delta R.T.: 0.002 min

Response: 37501070

Conc: 16.93 ng/ml