

Data Path : X:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL020619\
 Data File : PL044462.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Feb 2019 12:37
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 07 02:52:02 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020619.M
 Quant Title : GC Extractables
 QLast Update : Thu Feb 07 01:39:50 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.366	3.985	367.1E6	120.8E6	51.595	51.414
28)	SA Decachlor...	8.049	9.043	371.8E6	112.6E6	50.982	50.829
Target Compounds							
2)	A alpha-BHC	3.796	4.375	565.0E6	161.8E6	52.903	53.119
3)	MA gamma-BHC...	4.074	4.659	521.0E6	151.5E6	52.950	52.306
4)	MA Heptachlor	4.362	5.167	536.0E6	155.5E6	52.028	51.053
5)	MB Aldrin	4.603	5.471	510.3E6	143.5E6	53.010	51.786
6)	B beta-BHC	4.326	4.831	215.0E6	68660860	50.888	51.819
7)	B delta-BHC	4.523	5.047	432.0E6	100.7E6	53.659	55.096
8)	B Heptachlor...	5.040	5.858	480.3E6	128.2E6	51.623	51.263
9)	A Endosulfan I	5.373	6.214	453.7E6	129.7E6	51.994	51.248
10)	B gamma-Chl...	5.264	6.091	500.8E6	141.9E6	51.892	51.132
11)	B alpha-Chl...	5.320	6.164	494.2E6	141.3E6	51.690	51.012
12)	B 4,4'-DDE	5.487	6.321	458.5E6	125.4E6	52.873	52.521
13)	MA Dieldrin	5.612	6.470	494.0E6	135.1E6	52.628	52.364
14)	MA Endrin	5.866	6.687	458.2E6	116.3E6	52.308	51.773
15)	B Endosulfa...	6.140	6.897	438.6E6	118.4E6	51.869	51.385
16)	A 4,4'-DDD	5.999	6.811	382.3E6	100.7E6	52.850	52.423
17)	MA 4,4'-DDT	6.234	7.109	392.2E6	103.7E6	54.696	52.851
18)	B Endrin al...	6.307	7.021	364.3E6	102.5E6	51.959	50.503
19)	B Endosulfa...	6.518	7.243	384.9E6	107.5E6	52.954	50.952
20)	A Methoxychlor	6.779	7.567	212.2E6	61532427	52.883	51.504
21)	B Endrin ke...	7.002	7.713	446.0E6	118.4E6	53.607	52.187

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : X:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL020619\
Data File : PL044462.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Feb 2019 12:37
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 07 02:52:02 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020619.M
Quant Title : GC Extractables
QLast Update : Thu Feb 07 01:39:50 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

