

Data Path : X:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL020619\
 Data File : PL044509.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Feb 2019 00:11
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 07 03:00:15 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020619.M
 Quant Title : GC Extractables
 QLast Update : Thu Feb 07 01:39:50 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.367	3.986	357.8E6	118.9E6	50.288	50.633
28)	SA Decachlor...	8.051	9.045	349.0E6	104.5E6	47.856	47.181
Target Compounds							
2)	A alpha-BHC	3.797	4.376	550.7E6	159.9E6	51.558	52.477
3)	MA gamma-BHC...	4.074	4.660	506.5E6	149.9E6	51.479	51.743
4)	MA Heptachlor	4.363	5.168	514.7E6	151.8E6	49.961	49.845
5)	MB Aldrin	4.603	5.472	499.0E6	141.9E6	51.842	51.195
6)	B beta-BHC	4.327	4.832	213.1E6	68347735	50.420	51.583
7)	B delta-BHC	4.523	5.048	427.4E6	112.4E6	53.096	61.482
8)	B Heptachlor...	5.041	5.859	470.5E6	135.2E6	50.564	54.073
9)	A Endosulfan I	5.373	6.215	442.0E6	127.6E6	50.650	50.431
10)	B gamma-Chl...	5.264	6.092	487.6E6	140.6E6	50.523	50.651
11)	B alpha-Chl...	5.321	6.165	478.4E6	138.9E6	50.035	50.137
12)	B 4,4'-DDE	5.488	6.323	444.6E6	123.5E6	51.271	51.717
13)	MA Dieldrin	5.613	6.472	479.6E6	133.2E6	51.100	51.647
14)	MA Endrin	5.867	6.689	423.4E6	109.3E6	48.334	48.665
15)	B Endosulfa...	6.141	6.899	426.3E6	119.2E6	50.420	51.756
16)	A 4,4'-DDD	5.999	6.813	392.2E6	106.6E6	54.221	55.509
17)	MA 4,4'-DDT	6.235	7.111	334.3E6	91487136	46.613	46.634
18)	B Endrin al...	6.308	7.023	352.5E6	102.3E6	50.278	50.402
19)	B Endosulfa...	6.519	7.245	375.6E6	105.5E6	51.673	49.986
20)	A Methoxychlor	6.780	7.569	191.0E6	55421108	47.607	46.388
21)	B Endrin ke...	7.002	7.715	452.9E6	118.3E6	54.438	52.133

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : X:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL020619\
Data File : PL044509.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Feb 2019 00:11
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 07 03:00:15 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020619.M
Quant Title : GC Extractables
QLast Update : Thu Feb 07 01:39:50 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

