

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL020620\
 Data File : PL056364.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Feb 2020 14:37
 Operator : AJ\MA
 Sample : L1408-11
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 SHORE-RD-PATH-BH-04-A

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 07 06:49:54 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012820.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 28 13:10:37 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.498	4.016	132.3E6	39788362	12.631	13.607
28)	SA Decachlor...	8.254	9.096	133.4E6	27017946	12.672	11.945
Target Compounds							
17)	MA 4,4'-DDT	6.423	0.000	8349650	0	0.764	N.D. #
23)	Chlordane-1	4.375	0.000	21100674	0	55.742	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL020620\
Data File : PL056364.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Feb 2020 14:37
Operator : AJ\MA
Sample : L1408-11
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
SHORE-RD-PATH-BH-04-A

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 07 06:49:54 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL012820.M
Quant Title : GC Extractables
QLast Update : Tue Jan 28 13:10:37 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.498 min
Delta R.T.: 0.000 min
Response: 132287927
Conc: 12.63 ng/ml

Instrument :
ECD_L
ClientSampleId :
SHORE-RD-PATH-BH-04-A




#1 Tetrachloro-m-xylene

R.T.: 4.016 min
Delta R.T.: 0.000 min
Response: 39788362
Conc: 13.61 ng/ml




#17 4,4'-DDT

R.T.: 6.423 min
Delta R.T.: 0.004 min
Response: 8349650
Conc: 0.76 ng/ml



#17 4,4'-DDT


R.T.: 0.000 min
Exp R.T.: 7.146 min
Response: 0
Conc: N.D.



#23 Chlordane-1


R.T.: 4.375 min
Delta R.T.: 0.002 min
Response: 21100674
Conc: 55.74 ng/ml

Instrument :
ECD_L
ClientSampleId :
SHORE-RD-PATH-BH-04-A




#23 Chlordane-1

R.T.: 0.000 min
Exp R.T.: 5.008 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.254 min
Delta R.T.: 0.001 min
Response: 133409082
Conc: 12.67 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.096 min
Delta R.T.: 0.000 min
Response: 27017946
Conc: 11.94 ng/ml