

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL020821\
 Data File : PL064665.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Feb 2021 18:01
 Operator : AR\AJ
 Sample : PSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 09 01:44:03 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020821.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 09 01:42:47 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.427	4.103	136.7E6	105.3E6	50.000	50.000
28)	SA Decachlor...	8.149	9.263	109.8E6	93392518	50.000	50.000
Target Compounds							
2)	A alpha-BHC	3.861	4.507	228.8E6	151.8E6	50.000	50.000
3)	MA gamma-BHC...	4.142	4.799	210.7E6	140.6E6	50.000	50.000
4)	MA Heptachlor	4.438	5.310	201.4E6	143.3E6	50.000	50.000
5)	MB Aldrin	4.683	5.620	184.4E6	120.5E6	50.000	50.000
6)	B beta-BHC	4.393	4.978	83156746	65242072	50.000	50.000
7)	B delta-BHC	4.594	5.199	201.2E6	130.1E6	50.000	50.000
8)	B Heptachlo...	5.124	6.015	170.6E6	111.1E6	50.000	50.000
9)	A Endosulfan I	5.460	6.378	156.2E6	103.2E6	50.000	50.000
10)	B gamma-Chl...	5.349	6.250	170.8E6	113.5E6	50.000	50.000
11)	B alpha-Chl...	5.407	6.323	169.0E6	113.1E6	50.000	50.000
12)	B 4,4'-DDE	5.573	6.479	154.7E6	100.8E6	50.000	50.000
13)	MA Dieldrin	5.701	6.638	161.4E6	105.3E6	50.000	50.000
14)	MA Endrin	5.956	6.860	146.8E6	88756244	50.000	50.000
15)	B Endosulfa...	6.229	7.075	140.3E6	91815881	50.000	50.000
16)	A 4,4'-DDD	6.086	6.979	125.7E6	84794316	50.000	50.000
17)	MA 4,4'-DDT	6.325	7.278	114.5E6	77584256	50.000	50.000
18)	B Endrin al...	6.397	7.202	118.0E6	79423187	50.000	50.000
19)	B Endosulfa...	6.610	7.425	137.8E6	88701692	50.000	50.000
20)	A Methoxychlor	6.873	7.740	65187772	45749779	50.000	50.000
21)	B Endrin ke...	7.096	7.904	142.7E6	94822344	50.000	50.000
22)	Mirex	7.277	8.352	102.7E6	69088792	50.000	50.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL020821\
Data File : PL064665.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Feb 2021 18:01
Operator : AR\AJ
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 09 01:44:03 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020821.M
Quant Title : GC Extractables
QLast Update : Tue Feb 09 01:42:47 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

