

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021021\
 Data File : PL064749.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Feb 2021 09:53
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 2/11/2021 8:58:26 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 11 05:22:34 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020821.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 09 04:41:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.433	4.102	153.3E6	121.6E6	54.495	56.297
28)	SA Decachlor...	8.159	9.261	116.4E6	94184917	51.768	49.587
Target Compounds							
2)	A alpha-BHC	3.867	4.504	255.6E6	176.5E6	55.117	58.287
3)	MA gamma-BHC...	4.148	4.796	222.2E6	166.5E6	51.391	58.609
4)	MA Heptachlor	4.445	5.308	217.0E6	154.3E6	52.933	52.081
5)	MB Aldrin	4.691	5.618	203.7E6	135.9E6	54.705	55.421
6)	B beta-BHC	4.400	4.975	93399742	69572123	54.646	53.398
7)	B delta-BHC	4.601	5.197	228.4E6	143.2E6	56.515	53.013
8)	B Heptachlo...	5.132	6.013	188.2E6	125.1E6	54.219	56.701
9)	A Endosulfan I	5.468	6.376	173.9E6	115.0E6	54.229	54.825
10)	B gamma-Chl...	5.358	6.249	189.2E6	125.7E6	53.969	54.398
11)	B alpha-Chl...	5.416	6.321	187.3E6	125.4E6	54.078	54.468
12)	B 4,4'-DDE	5.582	6.478	170.7E6	109.4E6	54.542	53.956
13)	MA Dieldrin	5.709	6.636	179.9E6	114.5E6	54.850	53.570
14)	MA Endrin	5.964	6.857	148.5E6	87468003	49.439	48.384
15)	B Endosulfa...	6.238	7.073	156.9E6	99669781	54.490	52.894
16)	A 4,4'-DDD	6.095	6.977	148.1E6	89790098	58.491m	51.794
17)	MA 4,4'-DDT	6.334	7.276	117.7E6	78202974	52.215	51.292
18)	B Endrin al...	6.406	7.198	130.5E6	86080256	53.522	52.863
19)	B Endosulfa...	6.619	7.422	153.3E6	95523410	54.063	53.080
20)	A Methoxychlor	6.882	7.737	66510299	45405105	51.365	50.229
21)	B Endrin ke...	7.105	7.901	163.9E6	103.0E6	57.131	53.854
22)	Mirex	7.286	8.351	114.9E6	74822835	54.397	53.829

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL021021\
Data File : PL064749.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Feb 2021 09:53
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
2/11/2021 8:58:26 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 11 05:22:34 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL020821.M
Quant Title : GC Extractables
QLast Update : Tue Feb 09 04:41:29 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

